

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
 Data File : VX040230.D
 Acq On : 14 Feb 2024 18:59
 Operator : JC/MD
 Sample : P1462-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20240762-39N-N-FIELD-BLANK-1

Quant Time: Feb 15 00:07:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	68400	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	115104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	44034	43.273	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.540%
35) Dibromofluoromethane	5.385	113	36475	45.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.760%
50) Toluene-d8	8.647	98	137131	45.606	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.220%#
62) 4-Bromofluorobenzene	11.079	95	55446	45.213	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.420%
Target Compounds						
16) Acetone	2.380	43	2574	6.159	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
Data File : VX040230.D
Acq On : 14 Feb 2024 18:59
Operator : JC/MD
Sample : P1462-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20240762-39N-N-FIELD-BLANK-1

Quant Time: Feb 15 00:07:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

