

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
 Data File : VX040239.D
 Acq On : 14 Feb 2024 22:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 15 00:11:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.565	0.560	0.9	83	0.00
3 P	Chloromethane	0.582	0.583	-0.2	91	0.00
4 C	Vinyl Chloride	0.615	0.610	0.8#	85	0.00
5 T	Bromomethane	0.347	0.425	-22.5	95	0.00
6 T	Chloroethane	0.388	0.389	-0.3	90	0.00
7 T	Trichlorofluoromethane	0.988	0.931	5.8	83	0.00
8 T	Diethyl Ether	0.371	0.387	-4.3	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.530	0.507	4.3	85	0.00
10 T	Methyl Iodide	0.589	0.681	-15.6	97	0.00
11 T	Tert butyl alcohol	0.145	0.150	-3.4	92	0.00
12 CM	1,1-Dichloroethene	0.510	0.500	2.0#	88	0.00
13 T	Acrolein	0.134	0.138	-3.0	97	0.00
14 T	Allyl chloride	0.886	0.874	1.4	88	0.00
15 T	Acrylonitrile	0.324	0.338	-4.3	90	0.00
16 T	Acetone	0.306	0.264	13.7	78	0.00
17 T	Carbon Disulfide	1.348	1.200	11.0	83	0.00
18 T	Methyl Acetate	0.819	0.735	10.3	79	0.00
19 T	Methyl tert-butyl Ether	1.798	1.944	-8.1	93	0.00
20 T	Methylene Chloride	0.624	0.623	0.2	94	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.555	4.5	89	0.00
22 T	Diisopropyl ether	1.797	1.910	-6.3	92	0.00
23 T	Vinyl Acetate	1.602	1.763	-10.0	94	0.00
24 P	1,1-Dichloroethane	1.047	1.062	-1.4	90	0.00
25 T	2-Butanone	0.478	0.465	2.7	86	0.00
26 T	2,2-Dichloropropane	0.806	0.699	13.3	75	0.00
27 T	cis-1,2-Dichloroethene	0.663	0.676	-2.0	92	0.00
28 T	Bromochloromethane	0.444	0.480	-8.1	92	0.00
29 T	Tetrahydrofuran	0.311	0.316	-1.6	89	0.00
30 C	Chloroform	1.132	1.161	-2.6#	91	0.00
31 T	Cyclohexane	0.887	0.789	11.0	81	0.00
32 T	1,1,1-Trichloroethane	0.952	0.944	0.8	86	0.00
33 S	1,2-Dichloroethane-d4	0.744	0.779	-4.7	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.349	0.375	-7.4	98	0.00
36 T	1,1-Dichloropropene	0.489	0.448	8.4	82	0.00
37 T	Ethyl Acetate	0.544	0.555	-2.0	93	0.00
38 T	Carbon Tetrachloride	0.465	0.446	4.1	84	0.00
39 T	Methylcyclohexane	0.568	0.512	9.9	80	0.00
40 TM	Benzene	1.368	1.385	-1.2	89	0.00
41 T	Methacrylonitrile	0.289	0.306	-5.9	93	-0.01
42 TM	1,2-Dichloroethane	0.535	0.552	-3.2	88	0.00
43 T	Isopropyl Acetate	0.850	0.915	-7.6	94	0.00
44 TM	Trichloroethene	0.377	0.367	2.7	87	0.00
45 C	1,2-Dichloropropane	0.356	0.384	-7.9#	93	0.00
46 T	Dibromomethane	0.262	0.285	-8.8	92	0.00
47 T	Bromodichloromethane	0.505	0.538	-6.5	92	0.00
48 T	Methyl methacrylate	0.435	0.452	-3.9	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
 Data File : VX040239.D
 Acq On : 14 Feb 2024 22:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 15 00:11:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	88	0.00
50 S	Toluene-d8	1.306	1.331	-1.9	92	0.00
51 T	4-Methyl-2-Pentanone	0.580	0.619	-6.7	91	0.00
52 CM	Toluene	0.896	0.912	-1.8#	88	0.00
53 T	t-1,3-Dichloropropene	0.528	0.564	-6.8	88	0.00
54 T	cis-1,3-Dichloropropene	0.566	0.595	-5.1	88	0.00
55 T	1,1,2-Trichloroethane	0.382	0.406	-6.3	95	0.00
56 T	Ethyl methacrylate	0.572	0.635	-11.0	92	0.00
57 T	1,3-Dichloropropane	0.624	0.667	-6.9	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.271	0.286	-5.5	90	0.00
59 T	2-Hexanone	0.460	0.478	-3.9	88	0.00
60 T	Dibromochloromethane	0.389	0.426	-9.5	91	0.00
61 T	1,2-Dibromoethane	0.406	0.436	-7.4	93	0.00
62 S	4-Bromofluorobenzene	0.533	0.537	-0.8	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.350	0.337	3.7	90	0.00
65 PM	Chlorobenzene	1.044	1.017	2.6	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.379	-5.0	89	0.00
67 C	Ethyl Benzene	1.859	1.858	0.1#	86	0.00
68 T	m/p-Xylenes	0.702	0.697	0.7	86	0.00
69 T	o-Xylene	0.682	0.697	-2.2	87	0.00
70 T	Styrene	1.134	1.186	-4.6	88	0.00
71 P	Bromoform	0.296	0.322	-8.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.522	3.333	5.4	84	0.00
74 T	N-amyl acetate	1.510	1.684	-11.5	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.265	1.300	-2.8	92	0.00
76 T	1,2,3-Trichloropropane	1.013	1.087	-7.3	93	0.00
77 T	Bromobenzene	0.832	0.838	-0.7	90	0.00
78 T	n-propylbenzene	4.264	4.078	4.4	84	0.00
79 T	2-Chlorotoluene	2.575	2.407	6.5	84	0.00
80 T	1,3,5-Trimethylbenzene	2.965	2.905	2.0	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.328	0.338	-3.0	89	0.00
82 T	4-Chlorotoluene	3.050	2.840	6.9	84	0.00
83 T	tert-Butylbenzene	2.907	2.782	4.3	85	0.00
84 T	1,2,4-Trimethylbenzene	3.000	2.970	1.0	87	0.00
85 T	sec-Butylbenzene	3.736	3.570	4.4	84	0.00
86 T	p-Isopropyltoluene	3.062	2.962	3.3	83	0.00
87 T	1,3-Dichlorobenzene	1.648	1.617	1.9	87	0.00
88 T	1,4-Dichlorobenzene	1.745	1.656	5.1	87	0.00
89 T	n-Butylbenzene	2.956	2.763	6.5	82	0.00
90 T	Hexachloroethane	0.483	0.484	-0.2	84	0.00
91 T	1,2-Dichlorobenzene	1.595	1.596	-0.1	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.292	0.309	-5.8	89	0.00
93 T	1,2,4-Trichlorobenzene	1.092	1.058	3.1	88	0.00
94 T	Hexachlorobutadiene	0.436	0.405	7.1	87	0.00
95 T	Naphthalene	3.450	3.593	-4.1	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
Data File : VX040239.D
Acq On : 14 Feb 2024 22:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 15 00:11:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.041	1.036	0.5	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6