

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044952.D
 Acq On : 14 Feb 2025 11:40
 Operator : JC/MD
 Sample : VX0214WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0214WBSD01

Quant Time: Feb 14 22:12:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	103875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	184991	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	164889	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	77784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	77707	51.105	ug/l	0.00
Spiked Amount 50.000	Range	74 - 125	Recovery	=	102.200%	
35) Dibromofluoromethane	5.379	113	60838	50.582	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	101.160%	
50) Toluene-d8	8.647	98	226624	49.828	ug/l	0.00
Spiked Amount 50.000	Range	86 - 113	Recovery	=	99.660%	
62) 4-Bromofluorobenzene	11.079	95	81007	52.833	ug/l	0.00
Spiked Amount 50.000	Range	77 - 121	Recovery	=	105.660%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	27927	19.169	ug/l	99
3) Chloromethane	1.301	50	32990	18.605	ug/l	96
4) Vinyl Chloride	1.374	62	30252	17.611	ug/l	100
5) Bromomethane	1.593	94	11122	21.636	ug/l	97
6) Chloroethane	1.666	64	16528	22.848	ug/l	100
7) Trichlorofluoromethane	1.874	101	42055	19.359	ug/l	98
8) Diethyl Ether	2.130	74	15615	19.820	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	26009	19.816	ug/l	99
10) Methyl Iodide	2.447	142	32276	18.974	ug/l	97
11) Tert butyl alcohol	2.971	59	29076	103.075	ug/l	99
12) 1,1-Dichloroethene	2.313	96	24273	18.130	ug/l	97
13) Acrolein	2.233	56	31135	106.978	ug/l	98
14) Allyl chloride	2.660	41	45037	18.613	ug/l	98
15) Acrylonitrile	3.062	53	78868	104.716	ug/l	99
16) Acetone	2.380	43	64564	105.623	ug/l	96
17) Carbon Disulfide	2.508	76	60307	16.426	ug/l	100
18) Methyl Acetate	2.697	43	41481	21.504	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	84252	19.778	ug/l	99
20) Methylene Chloride	2.782	84	28525	19.052	ug/l	93
21) trans-1,2-Dichloroethene	3.087	96	24497	18.590	ug/l	100
22) Diisopropyl ether	3.757	45	91929	20.082	ug/l	94
23) Vinyl Acetate	3.715	43	372164	99.722	ug/l	99
24) 1,1-Dichloroethane	3.605	63	50075	19.550	ug/l	99
25) 2-Butanone	4.556	43	105945	106.736	ug/l	97
26) 2,2-Dichloropropane	4.465	77	37106	18.267	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	30448	19.237	ug/l	99
28) Bromochloromethane	4.891	49	22191	18.027	ug/l	100
29) Tetrahydrofuran	5.001	42	68840	104.445	ug/l	99
30) Chloroform	5.093	83	49641	19.984	ug/l	96
31) Cyclohexane	5.458	56	44062	18.660	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	40495	19.219	ug/l	98
36) 1,1-Dichloropropene	5.684	75	32376	18.658	ug/l	99
37) Ethyl Acetate	4.715	43	40204	20.830	ug/l	99
38) Carbon Tetrachloride	5.666	117	33498	19.700	ug/l	93
39) Methylcyclohexane	7.373	83	44129	19.671	ug/l	98
40) Benzene	6.031	78	106024	19.567	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044952.D
 Acq On : 14 Feb 2025 11:40
 Operator : JC/MD
 Sample : VX0214WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0214WBSD01

Quant Time: Feb 14 22:12:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	23036	22.298	ug/l	98
42) 1,2-Dichloroethane	6.080	62	37652	21.803	ug/l	100
43) Isopropyl Acetate	6.336	43	64357	20.648	ug/l	98
44) Trichloroethene	7.123	130	24052	19.420	ug/l	96
45) 1,2-Dichloropropane	7.427	63	26311	19.518	ug/l	99
46) Dibromomethane	7.580	93	19515	20.875	ug/l	98
47) Bromodichloromethane	7.818	83	37748	20.853	ug/l	97
48) Methyl methacrylate	7.690	41	31848	21.671	ug/l	97
49) 1,4-Dioxane	7.659	88	13982	443.448	ug/l	98
51) 4-Methyl-2-Pentanone	8.567	43	219588	115.888	ug/l	97
52) Toluene	8.714	92	63830	19.780	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	35347	19.296	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	40620	19.889	ug/l	99
55) 1,1,2-Trichloroethane	9.147	97	26556	21.408	ug/l	98
56) Ethyl methacrylate	9.116	69	41272	20.600	ug/l	99
57) 1,3-Dichloropropane	9.305	76	45713	21.061	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	99532	101.176	ug/l	99
59) 2-Hexanone	9.427	43	159284	116.671	ug/l	99
60) Dibromochloromethane	9.519	129	28129	21.203	ug/l	97
61) 1,2-Dibromoethane	9.604	107	27025	21.494	ug/l	97
64) Tetrachloroethene	9.269	164	21103	20.294	ug/l	96
65) Chlorobenzene	10.073	112	70853	20.006	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	22743	19.951	ug/l	98
67) Ethyl Benzene	10.189	91	122687	19.631	ug/l	100
68) m/p-Xylenes	10.299	106	92530	40.136	ug/l	99
69) o-Xylene	10.640	106	45891	19.838	ug/l	97
70) Styrene	10.652	104	77302	20.741	ug/l	99
71) Bromoform	10.799	173	18090	21.264	ug/l #	100
73) Isopropylbenzene	10.957	105	117063	18.691	ug/l	99
74) N-amyl acetate	10.841	43	55559	19.604	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.207	83	42130	19.576	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	42262	24.506	ug/l	85
77) Bromobenzene	11.195	156	27335	19.251	ug/l	99
78) n-propylbenzene	11.299	91	138085	19.101	ug/l	100
79) 2-Chlorotoluene	11.360	91	84337	18.982	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	98051	19.292	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	11059	16.496	ug/l	89
82) 4-Chlorotoluene	11.451	91	94538	19.233	ug/l	99
83) tert-Butylbenzene	11.713	119	99538	19.359	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	100112	19.957	ug/l	100
85) sec-Butylbenzene	11.890	105	126113	19.808	ug/l	99
86) p-Isopropyltoluene	12.006	119	103751	20.263	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	50550	19.359	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	51205	19.395	ug/l	99
89) n-Butylbenzene	12.329	91	90481	19.975	ug/l	99
90) Hexachloroethane	12.536	117	17246	17.989	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	52530	20.469	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	7941	20.217	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	29608	19.066	ug/l	98
94) Hexachlorobutadiene	13.725	225	12223	19.827	ug/l	98
95) Naphthalene	13.774	128	110366	19.661	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	30613	19.564	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044952.D
Acq On : 14 Feb 2025 11:40
Operator : JC/MD
Sample : VX0214WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01

Quant Time: Feb 14 22:12:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044952.D
Acq On : 14 Feb 2025 11:40
Operator : JC/MD
Sample : VX0214WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01

Quant Time: Feb 14 22:12:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. 0.001 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:168 Resp: 103875
Ion Ratio Lower Upper
168 100
99 61.3 51.2 76.8



#2
Dichlorodifluoromethane
Concen: 19.169 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 85 Resp: 27927
Ion Ratio Lower Upper
85 100
87 32.3 15.8 47.3





#3

Chloromethane

Concen: 18.605 ug/l

RT: 1.301 min Scan# 35

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :





VX0214WBSD01

Tgt Ion: 50 Resp: 32990

Ion Ratio Lower Upper

50 100

52 30.4 25.9 38.9



#4

Vinyl Chloride

Concen: 17.611 ug/l

RT: 1.374 min Scan# 47

Delta R.T. 0.000 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion: 62 Resp: 30252

Ion Ratio Lower Upper

62 100

64 31.6 25.4 38.0









#5
Bromomethane
Concen: 21.636 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 94 Resp: 11122
Ion Ratio Lower Upper
94 100
96 92.1 71.6 107.4



#6
Chloroethane
Concen: 22.848 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 64 Resp: 16528
Ion Ratio Lower Upper
64 100
66 32.5 25.9 38.9









#7
Trichlorofluoromethane
Concen: 19.359 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.001 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:101 Resp: 42055
Ion Ratio Lower Upper
101 100
103 63.7 52.4 78.6



#8
Diethyl Ether
Concen: 19.820 ug/l
RT: 2.130 min Scan# 171
Delta R.T. 0.001 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 74 Resp: 15615
Ion Ratio Lower Upper
74 100
45 103.6 51.1 153.4









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.816 ug/l
RT: 2.325 min Scan# 20
Delta R.T. 0.006 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:101 Resp: 26009
Ion Ratio Lower Upper
101 100
85 44.5 35.7 53.5
151 68.9 56.0 84.0



#10
Methyl Iodide
Concen: 18.974 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:142 Resp: 32276
Ion Ratio Lower Upper
142 100
127 45.7 34.3 51.5
141 14.3 11.4 17.2









#11
Tert butyl alcohol
Concen: 103.075 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.012 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 59 Resp: 29076
Ion Ratio Lower Upper
59 100
57 9.8 8.1 12.1



#12
1,1-Dichloroethene
Concen: 18.130 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 96 Resp: 24273
Ion Ratio Lower Upper
96 100
61 168.8 130.2 195.4
98 65.9 52.2 78.4









#13
Acrolein
Concen: 106.978 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 56 Resp: 31135
Ion Ratio Lower Upper
56 100
55 71.0 55.7 83.5



#14
Allyl chloride
Concen: 18.613 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 41 Resp: 45037
Ion Ratio Lower Upper
41 100
39 66.8 51.4 77.2
76 33.2 27.0 40.4





#15

Acrylonitrile

Concen: 104.716 ug/l

RT: 3.062 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion: 53 Resp: 78868

Ion Ratio Lower Upper

53 100

52 81.7 65.0 97.6

51 36.2 28.0 42.0



#16

Acetone

Concen: 105.623 ug/l

RT: 2.380 min Scan# 212

Delta R.T. 0.000 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion: 43 Resp: 64564

Ion Ratio Lower Upper

43 100

58 30.7 26.5 39.7









#17
Carbon Disulfide
Concen: 16.426 ug/l
RT: 2.508 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 76 Resp: 60307
Ion Ratio Lower Upper
76 100
78 8.7 7.0 10.4



#18
Methyl Acetate
Concen: 21.504 ug/l
RT: 2.697 min Scan# 264
Delta R.T. -0.006 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 43 Resp: 41481
Ion Ratio Lower Upper
43 100
74 22.4 18.9 28.3









#19
Methyl tert-butyl Ether
Concen: 19.778 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 73 Resp: 84252
Ion Ratio Lower Upper
73 100
57 21.8 17.9 26.9



#20
Methylene Chloride
Concen: 19.052 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 84 Resp: 28525
Ion Ratio Lower Upper
84 100
49 135.3 100.1 150.1
51 42.2 31.7 47.5
86 68.0 50.8 76.2









#21
trans-1,2-Dichloroethene
Concen: 18.590 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 96 Resp: 24497
Ion Ratio Lower Upper
96 100
61 132.3 106.0 159.0
98 62.5 50.6 76.0



#22
Diisopropyl ether
Concen: 20.082 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 45 Resp: 91929
Ion Ratio Lower Upper
45 100
43 78.9 58.0 87.0
87 26.6 20.7 31.1
59 12.0 9.2 13.8









#23
 Vinyl Acetate
 Concen: 99.722 ug/l
 RT: 3.715 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX044952.D
 Acq: 14 Feb 2025 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0214WBSD01


Tgt Ion: 43 Resp: 372164
 Ion Ratio Lower Upper
 43 100
 86 9.9 8.3 12.5



#24
 1,1-Dichloroethane
 Concen: 19.550 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX044952.D
 Acq: 14 Feb 2025 11:40

Tgt Ion: 63 Resp: 50075
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.3 9.9
 100 3.8 2.1 6.5









#25
2-Butanone
Concen: 106.736 ug/l
RT: 4.556 min Scan# 509
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 43 Resp: 105945
Ion Ratio Lower Upper
43 100
72 25.0 21.2 31.8



#26
2,2-Dichloropropane
Concen: 18.267 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 77 Resp: 37106
Ion Ratio Lower Upper
77 100
97 23.2 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 19.237 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion: 96 Resp: 30448

Ion Ratio Lower Upper

96 100

61 146.8 0.0 294.0

98 64.0 0.0 131.2



#28

Bromochloromethane

Concen: 18.027 ug/l

RT: 4.891 min Scan# 624

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40




Tgt Ion: 49 Resp: 22191


Ion Ratio Lower Upper

49 100

129 1.7 0.0 3.6

130 71.1 56.7 85.1





#29

Tetrahydrofuran

Concen: 104.445 ug/l

RT: 5.001 min Scan# 64

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion: 42 Resp: 68840

Ion Ratio Lower Upper

42 100

72 44.4 35.0 52.4

71 40.0 32.6 48.8



#30

Chloroform

Concen: 19.984 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion: 83 Resp: 49641

Ion Ratio Lower Upper

83 100

85 61.9 52.1 78.1









#31
Cyclohexane
Concen: 18.660 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 56 Resp: 44062
Ion Ratio Lower Upper
56 100
69 28.3 24.4 36.6
84 85.1 69.0 103.6



#32
1,1,1-Trichloroethane
Concen: 19.219 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 97 Resp: 40495
Ion Ratio Lower Upper
97 100
99 65.3 51.0 76.6
61 46.9 36.6 55.0





#33

1,2-Dichloroethane-d4

Concen: 51.105 ug/l

RT: 5.946 min Scan# 797

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :





VX0214WBSD01

Tgt Ion: 65 Resp: 77707

Ion Ratio Lower Upper

65 100

67 52.8 0.0 108.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40




Tgt Ion: 114 Resp: 184991


Ion Ratio Lower Upper

114 100

63 21.4 0.0 42.4


88 16.6 0.0 31.4








#35
Dibromofluoromethane
Concen: 50.582 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01







Tgt Ion:113 Resp: 60838
Ion Ratio Lower Upper
113 100
111 103.6 83.5 125.3
192 18.1 14.4 21.6



#36
1,1-Dichloropropene
Concen: 18.658 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 75 Resp: 32376
Ion Ratio Lower Upper
75 100
110 34.6 17.2 51.6
77 31.9 24.6 37.0





#37

Ethyl Acetate

Concen: 20.830 ug/l

RT: 4.715 min Scan# 594

Delta R.T. 0.006 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion: 43 Resp: 40204

Ion Ratio Lower Upper

43 100

61 13.1 11.0 16.4

70 9.7 8.2 12.2



#38

Carbon Tetrachloride

Concen: 19.700 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40




Tgt Ion: 117 Resp: 33498


Ion Ratio Lower Upper

117 100

119 89.5 77.5 116.3

121 29.2 25.6 38.4









#39
Methylcyclohexane
Concen: 19.671 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 83 Resp: 44129
Ion Ratio Lower Upper
83 100
55 75.7 61.3 91.9
98 48.1 37.0 55.6



#40
Benzene
Concen: 19.567 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 78 Resp: 106024
Ion Ratio Lower Upper
78 100
77 24.9 19.0 28.6









#41
Methacrylonitrile
Concen: 22.298 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:	41	Resp:	23036
Ion	Ratio	Lower	Upper
41	100		
39	55.0	43.1	64.7
67	69.6	57.0	85.6
52	30.4	24.0	36.0



#42
1,2-Dichloroethane
Concen: 21.803 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:	62	Resp:	37652
Ion	Ratio	Lower	Upper
62	100		
98	9.7	0.0	19.4









#43
Isopropyl Acetate
Concen: 20.648 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 43 Resp: 64357
Ion Ratio Lower Upper
43 100
61 18.6 15.8 23.8
87 12.3 10.0 15.0



#44
Trichloroethene
Concen: 19.420 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:130 Resp: 24052
Ion Ratio Lower Upper
130 100
95 97.4 0.0 202.4





#45

1,2-Dichloropropane

Concen: 19.518 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :





VX0214WBSD01

Tgt Ion: 63 Resp: 26311

Ion Ratio Lower Upper

63 100

65 29.9 24.2 36.4



#46

Dibromomethane

Concen: 20.875 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40




Tgt Ion: 93 Resp: 19515


Ion Ratio Lower Upper

93 100

95 82.2 66.2 99.4

174 90.0 69.9 104.9









#47
Bromodichloromethane
Concen: 20.853 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:	83	Resp:	37748
Ion	Ratio	Lower	Upper
83	100		
85	65.3	50.0	75.0
127	7.7	6.4	9.6



#48
Methyl methacrylate
Concen: 21.671 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:	41	Resp:	31848
Ion	Ratio	Lower	Upper
41	100		
69	79.7	65.8	98.8
39	60.1	46.4	69.6





#49

1,4-Dioxane

Concen: 443.448 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion: 88 Resp: 13982

Ion Ratio Lower Upper

88 100

43 35.9 26.6 39.8

58 69.9 55.9 83.9



#50

Toluene-d8

Concen: 49.828 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion: 98 Resp: 226624

Ion Ratio Lower Upper

98 100

100 64.2 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 115.888 ug/l
RT: 8.567 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 43 Resp: 219588
Ion Ratio Lower Upper
43 100
58 36.9 30.8 46.2



#52
Toluene
Concen: 19.780 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 92 Resp: 63830
Ion Ratio Lower Upper
92 100
91 176.5 137.4 206.2





#53

t-1,3-Dichloropropene

Concen: 19.296 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :





VX0214WBSD01

Tgt Ion: 75 Resp: 35347

Ion Ratio Lower Upper

75 100

77 32.7 26.0 39.0



#54

cis-1,3-Dichloropropene

Concen: 19.889 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40




Tgt Ion: 75 Resp: 40620


Ion Ratio Lower Upper

75 100

77 32.1 25.3 37.9

39 50.6 40.0 60.0









#55
1,1,2-Trichloroethane
Concen: 21.408 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:	97	Resp:	26556
Ion	Ratio	Lower	Upper
97	100		
83	88.8	69.8	104.6
85	55.0	44.8	67.2
99	61.3	50.6	76.0



#56
Ethyl methacrylate
Concen: 20.600 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:	69	Resp:	41272
Ion	Ratio	Lower	Upper
69	100		
41	69.9	55.2	82.8
39	38.8	30.8	46.2





#57

1,3-Dichloropropane

Concen: 21.061 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :





VX0214WBSD01

Tgt Ion: 76 Resp: 45713

Ion Ratio Lower Upper

76 100

78 33.5 25.8 38.8



#58

2-Chloroethyl Vinyl ether

Concen: 101.176 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion: 63 Resp: 99532

Ion Ratio Lower Upper

63 100

106 27.0 21.3 31.9









#59
2-Hexanone
Concen: 116.671 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 43 Resp: 159284
Ion Ratio Lower Upper
43 100
58 54.1 27.5 82.4



#60
Dibromochloromethane
Concen: 21.203 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 129 Resp: 28129
Ion Ratio Lower Upper
129 100
127 75.6 39.1 117.3









#61
1,2-Dibromoethane
Concen: 21.494 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:107 Resp: 27025
Ion Ratio Lower Upper
107 100
109 92.2 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 52.833 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 95 Resp: 81007
Ion Ratio Lower Upper
95 100
174 73.6 0.0 142.6
176 70.3 0.0 141.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:117 Resp: 164889
Ion Ratio Lower Upper
117 100
82 60.0 47.5 71.3
119 31.6 25.4 38.0



#64
Tetrachloroethene
Concen: 20.294 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:164 Resp: 21103
Ion Ratio Lower Upper
164 100
166 130.8 100.4 150.6
129 103.4 79.9 119.9
131 97.7 75.9 113.9









#65
Chlorobenzene
Concen: 20.006 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:112 Resp: 70853
Ion Ratio Lower Upper
112 100
114 32.0 25.0 37.6



#66
1,1,1,2-Tetrachloroethane
Concen: 19.951 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:131 Resp: 22743
Ion Ratio Lower Upper
131 100
133 97.0 47.1 141.3
119 67.9 33.6 100.6









#67
Ethyl Benzene
Concen: 19.631 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion: 91 Resp: 122687
Ion Ratio Lower Upper
91 100
106 29.7 23.9 35.9



#68
m/p-Xylenes
Concen: 40.136 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 106 Resp: 92530
Ion Ratio Lower Upper
106 100
91 207.2 164.3 246.5









#69
o-Xylene
Concen: 19.838 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:106 Resp: 45891
Ion Ratio Lower Upper
106 100
91 221.9 108.5 325.5



#70
Styrene
Concen: 20.741 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:104 Resp: 77302
Ion Ratio Lower Upper
104 100
78 51.2 40.5 60.7
103 54.7 43.6 65.4





#71

Bromoform

Concen: 21.264 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion:173 Resp: 18090

Ion Ratio Lower Upper

173 100

175 48.6 24.3 72.8

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40




Tgt Ion:152 Resp: 77784


Ion Ratio Lower Upper

152 100

115 71.7 43.4 130.1

150 160.7 0.0 350.4









#73
Isopropylbenzene
Concen: 18.691 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:105 Resp: 117063
Ion Ratio Lower Upper
105 100
120 26.0 13.2 39.5



#74
N-ethyl acetate
Concen: 19.604 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 43 Resp: 55559
Ion Ratio Lower Upper
43 100
70 40.6 33.7 50.5
55 22.9 18.6 27.8
61 22.8 19.1 28.7





#75

1,1,2,2-Tetrachloroethane

Concen: 19.576 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion: 83 Resp: 42130

Ion Ratio Lower Upper

83 100

131 9.0 4.5 13.5

85 64.4 32.3 96.8



#76

1,2,3-Trichloropropane

Concen: 24.506 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion: 75 Resp: 42262

Ion Ratio Lower Upper

75 100

77 34.1 21.9 65.7









#77
Bromobenzene
Concen: 19.251 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:156 Resp: 27335
Ion Ratio Lower Upper
156 100
77 175.6 88.4 265.3
158 97.2 48.4 145.2



#78
n-propylbenzene
Concen: 19.101 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 91 Resp: 138085
Ion Ratio Lower Upper
91 100
120 22.3 11.1 33.3





#79

2-Chlorotoluene

Concen: 18.982 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :





VX0214WBSD01

Tgt Ion: 91 Resp: 84337

Ion Ratio Lower Upper

91 100

126 32.8 16.2 48.6



#80

1,3,5-Trimethylbenzene

Concen: 19.292 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion: 105 Resp: 98051

Ion Ratio Lower Upper

105 100

120 48.9 24.0 72.0









#81
 trans-1,4-Dichloro-2-butene
 Concen: 16.496 ug/l
 RT: 11.018 min Scan# 1629
 Delta R.T. 0.000 min
 Lab File: VX044952.D
 Acq: 14 Feb 2025 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0214WBSD01


Tgt Ion: 75 Resp: 11059
 Ion Ratio Lower Upper
 75 100
 53 113.0 78.0 117.0
 89 44.6 36.6 54.8



#82
 4-Chlorotoluene
 Concen: 19.233 ug/l
 RT: 11.451 min Scan# 1700
 Delta R.T. 0.000 min
 Lab File: VX044952.D
 Acq: 14 Feb 2025 11:40

Tgt Ion: 91 Resp: 94538
 Ion Ratio Lower Upper
 91 100
 126 29.1 14.2 42.6





#83

tert-Butylbenzene

Concen: 19.359 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion:119 Resp: 99538

Ion Ratio Lower Upper

119 100

91 59.2 29.1 87.4

134 23.7 11.5 34.5



#84

1,2,4-Trimethylbenzene

Concen: 19.957 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion:105 Resp: 100112

Ion Ratio Lower Upper

105 100

120 44.6 22.4 67.0









#85
 sec-Butylbenzene
 Concen: 19.808 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX044952.D
 Acq: 14 Feb 2025 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0214WBSD01


Tgt Ion:105 Resp: 126113
 Ion Ratio Lower Upper
 105 100
 134 19.0 9.8 29.3



#86
 p-Isopropyltoluene
 Concen: 20.263 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX044952.D
 Acq: 14 Feb 2025 11:40

Tgt Ion:119 Resp: 103751
 Ion Ratio Lower Upper
 119 100
 134 25.3 13.0 38.9
 91 24.3 12.1 36.3









#87
1,3-Dichlorobenzene
Concen: 19.359 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:146 Resp: 50550
Ion Ratio Lower Upper
146 100
111 41.5 21.2 63.6
148 62.9 32.5 97.5



#88
1,4-Dichlorobenzene
Concen: 19.395 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:146 Resp: 51205
Ion Ratio Lower Upper
146 100
111 42.2 20.4 61.3
148 64.7 32.0 96.0





#89

n-Butylbenzene

Concen: 19.975 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX044952.D

Acq: 14 Feb 2025 11:40

Instrument :

MSVOA_X

ClientSampleId :

VX0214WBSD01





Tgt Ion: 91 Resp: 90481

Ion Ratio Lower Upper

91 100

92 53.9 27.2 81.6

134 23.3 12.1 36.3



#90

Hexachloroethane

Concen: 17.989 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX044952.D




Acq: 14 Feb 2025 11:40


Tgt Ion: 117 Resp: 17246

Ion Ratio Lower Upper

117 100

201 63.8 32.5 97.5









#91
1,2-Dichlorobenzene
Concen: 20.469 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:146 Resp: 52530
Ion Ratio Lower Upper
146 100
111 43.5 21.8 65.4
148 62.6 31.9 95.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.217 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion: 75 Resp: 7941
Ion Ratio Lower Upper
75 100
155 76.6 40.5 121.4
157 96.7 51.4 154.3









#93
1,2,4-Trichlorobenzene
Concen: 19.066 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40




Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01


Tgt Ion:180 Resp: 29608
Ion Ratio Lower Upper
180 100
182 97.1 47.6 142.8
145 32.3 16.4 49.1



#94
Hexachlorobutadiene
Concen: 19.827 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:225 Resp: 12223
Ion Ratio Lower Upper
225 100
223 62.4 31.4 94.2
227 61.6 32.3 96.8









#95
Naphthalene
Concen: 19.661 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Instrument :
MSVOA_X
ClientSampleId :
VX0214WBSD01

Tgt Ion:128 Resp: 110366
Ion Ratio Lower Upper
128 100
127 12.6 10.1 15.1
129 10.5 8.9 13.3



#96
1,2,3-Trichlorobenzene
Concen: 19.564 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044952.D
Acq: 14 Feb 2025 11:40

Tgt Ion:180 Resp: 30613
Ion Ratio Lower Upper
180 100
182 95.0 47.9 143.6
145 35.5 17.1 51.3

