

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044954.D
 Acq On : 14 Feb 2025 12:26
 Operator : JC/MD
 Sample : Q1347-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-EB-20250207

Quant Time: Feb 14 22:14:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	171322	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	154790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64957	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	70435	57.115	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.240%
35) Dibromofluoromethane	5.379	113	58500	52.519	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.040%
50) Toluene-d8	8.647	98	209395	49.714	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.079	95	72716	51.210	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.420%
Target Compounds						Qvalue
16) Acetone	2.380	43	3820	7.705	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044954.D
Acq On : 14 Feb 2025 12:26
Operator : JC/MD
Sample : Q1347-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-EB-20250207

Quant Time: Feb 14 22:14:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

