

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044959.D
 Acq On : 14 Feb 2025 14:20
 Operator : JC/MD
 Sample : PB166701TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166701TB

Quant Time: Feb 14 22:14:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	79050	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	158335	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145238	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64501	55.741	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.480%
35) Dibromofluoromethane	5.379	113	53471	51.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.880%
50) Toluene-d8	8.647	98	199097	51.146	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.300%
62) 4-Bromofluorobenzene	11.079	95	70644	53.831	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.660%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	26548	57.070	ug/l	98
18) Methyl Acetate	2.709	43	1614	1.099	ug/l #	82
20) Methylene Chloride	2.782	84	1644	1.443	ug/l	90
43) Isopropyl Acetate	6.336	43	100208	37.564	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044959.D
Acq On : 14 Feb 2025 14:20
Operator : JC/MD
Sample : PB166701TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166701TB

Quant Time: Feb 14 22:14:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

