

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044963.D
 Acq On : 14 Feb 2025 15:52
 Operator : JC/MD
 Sample : Q1356-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-PILE

Quant Time: Feb 14 22:15:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	80113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164255	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	150904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67224	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66539	56.740	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.480%
35) Dibromofluoromethane	5.379	113	55269	51.753	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.500%
50) Toluene-d8	8.647	98	204924	50.745	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.500%
62) 4-Bromofluorobenzene	11.079	95	72007	52.892	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.780%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	13834	29.344	ug/l	98
18) Methyl Acetate	2.703	43	2515	1.691	ug/l	91
43) Isopropyl Acetate	6.336	43	96164	34.749	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044963.D
Acq On : 14 Feb 2025 15:52
Operator : JC/MD
Sample : Q1356-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-PILE

Quant Time: Feb 14 22:15:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

