

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044964.D
 Acq On : 14 Feb 2025 16:15
 Operator : JC/MD
 Sample : Q1363-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-30D-021125

Quant Time: Feb 14 22:16:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	78193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	161133	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	143961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	58167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63817	55.755	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.500%
35) Dibromofluoromethane	5.379	113	54021	51.565	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.120%
50) Toluene-d8	8.647	98	200150	50.523	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.040%
62) 4-Bromofluorobenzene	11.079	95	65634	49.145	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.300%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	16240	14.808	ug/l	97
7) Trichlorofluoromethane	1.880	101	1609	0.984	ug/l #	83
9) 1,1,2-Trichlorotrifluo...	2.325	101	9002	9.111	ug/l	97
16) Acetone	2.373	43	1036	2.251	ug/l	93
18) Methyl Acetate	2.703	43	1786	1.230	ug/l #	87
27) cis-1,2-Dichloroethene	4.489	96	6660	5.590	ug/l	92
32) 1,1,1-Trichloroethane	5.373	97	2421	1.526	ug/l	99
44) Trichloroethene	7.123	130	6814	6.316	ug/l	99
64) Tetrachloroethene	9.275	164	14128	15.561	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044964.D
Acq On : 14 Feb 2025 16:15
Operator : JC/MD
Sample : Q1363-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-30D-021125

Quant Time: Feb 14 22:16:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

