

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044965.D
 Acq On : 14 Feb 2025 16:38
 Operator : JC/MD
 Sample : Q1363-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-021125

Quant Time: Feb 14 22:16:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	169043	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	154488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68065	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67873	56.755	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.500%
35) Dibromofluoromethane	5.385	113	55795	50.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	8.646	98	207369	49.896	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.800%
62) 4-Bromofluorobenzene	11.079	95	73559	52.502	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.000%
Target Compounds						
16) Acetone	2.379	43	13153	27.359	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044965.D
Acq On : 14 Feb 2025 16:38
Operator : JC/MD
Sample : Q1363-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-021125

Quant Time: Feb 14 22:16:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

