

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044967.D
 Acq On : 14 Feb 2025 17:25
 Operator : JC/MD
 Sample : Q1363-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VFCC-2-021225

Quant Time: Feb 14 22:17:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	78469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	163329	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	153346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66539	57.928	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 115.860%	
35) Dibromofluoromethane	5.379	113	55153	51.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.880%	
50) Toluene-d8	8.647	98	205791	51.249	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.500%	
62) 4-Bromofluorobenzene	11.079	95	73777	54.500	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 109.000%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	772	0.779	ug/l	96
12) 1,1-Dichloroethene	2.325	96	386	0.382	ug/l #	79
16) Acetone	2.380	43	943	2.042	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	3763	3.147	ug/l	94
30) Chloroform	5.093	83	977	0.521	ug/l	98
44) Trichloroethene	7.129	130	4393	4.017	ug/l	100
64) Tetrachloroethene	9.269	164	523	0.541	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044967.D
Acq On : 14 Feb 2025 17:25
Operator : JC/MD
Sample : Q1363-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VFCC-2-021225

Quant Time: Feb 14 22:17:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

