

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044971.D
 Acq On : 14 Feb 2025 18:57
 Operator : JC/MD
 Sample : Q1363-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-33D-021225

Quant Time: Feb 14 22:19:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	75470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	157095	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64939	58.782	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 117.560%	
35) Dibromofluoromethane	5.385	113	53696	52.572	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.140%	
50) Toluene-d8	8.647	98	196972	50.999	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.000%	
62) 4-Bromofluorobenzene	11.079	95	70404	54.072	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 108.140%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	8582	8.108	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.331	101	2794	2.930	ug/l	92
16) Acetone	2.380	43	835	1.880	ug/l	94
27) cis-1,2-Dichloroethene	4.495	96	2340	2.035	ug/l	93
44) Trichloroethene	7.123	130	8038	7.643	ug/l	100
64) Tetrachloroethene	9.275	164	2414	2.594	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044971.D
Acq On : 14 Feb 2025 18:57
Operator : JC/MD
Sample : Q1363-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-33D-021225

Quant Time: Feb 14 22:19:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

