

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044948.D
 Acq On : 14 Feb 2025 10:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 22:10:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.701	0.746	-6.4	93	0.00
3 P	Chloromethane	0.854	0.855	-0.1	88	0.00
4 C	Vinyl Chloride	0.827	0.803	2.9#	87	0.00
5 T	Bromomethane	0.247	0.285	-15.4	100	0.00
6 T	Chloroethane	0.307	0.432	-40.7#	110	0.00
7 T	Trichlorofluoromethane	1.046	1.097	-4.9	93	0.00
8 T	Diethyl Ether	0.379	0.394	-4.0	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.656	-3.8	92	0.00
10 T	Methyl Iodide	0.819	0.868	-6.0	89	0.00
11 T	Tert butyl alcohol	0.136	0.134	1.5	85	-0.01
12 CM	1,1-Dichloroethene	0.644	0.637	1.1#	88	0.00
13 T	Acrolein	0.140	0.153	-9.3	94	0.00
14 T	Allyl chloride	1.165	1.167	-0.2	88	0.00
15 T	Acrylonitrile	0.363	0.378	-4.1	87	0.00
16 T	Acetone	0.294	0.310	-5.4	92	0.00
17 T	Carbon Disulfide	1.767	1.631	7.7	81	0.00
18 T	Methyl Acetate	0.928	0.982	-5.8	91	0.00
19 T	Methyl tert-butyl Ether	2.050	2.063	-0.6	88	0.00
20 T	Methylene Chloride	0.721	0.709	1.7	88	0.00
21 T	trans-1,2-Dichloroethene	0.634	0.627	1.1	85	0.00
22 T	Diisopropyl ether	2.203	2.293	-4.1	89	0.00
23 T	Vinyl Acetate	1.796	1.866	-3.9	88	0.00
24 P	1,1-Dichloroethane	1.233	1.248	-1.2	89	0.00
25 T	2-Butanone	0.478	0.507	-6.1	87	0.00
26 T	2,2-Dichloropropane	0.978	0.973	0.5	89	-0.01
27 T	cis-1,2-Dichloroethene	0.762	0.762	0.0	88	0.00
28 T	Bromochloromethane	0.593	0.594	-0.2	89	0.00
29 T	Tetrahydrofuran	0.317	0.328	-3.5	87	-0.01
30 C	Chloroform	1.196	1.212	-1.3#	90	0.00
31 T	Cyclohexane	1.137	1.103	3.0	86	0.00
32 T	1,1,1-Trichloroethane	1.014	1.027	-1.3	89	0.00
33 S	1,2-Dichloroethane-d4	0.732	0.726	0.8	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.325	0.335	-3.1	87	0.00
36 T	1,1-Dichloropropene	0.469	0.475	-1.3	86	0.00
37 T	Ethyl Acetate	0.522	0.551	-5.6	87	0.00
38 T	Carbon Tetrachloride	0.460	0.493	-7.2	91	0.00
39 T	Methylcyclohexane	0.606	0.650	-7.3	87	0.00
40 TM	Benzene	1.465	1.515	-3.4	86	0.00
41 T	Methacrylonitrile	0.279	0.321	-15.1	89	0.00
42 TM	1,2-Dichloroethane	0.467	0.510	-9.2	90	0.00
43 T	Isopropyl Acetate	0.842	0.914	-8.6	87	0.00
44 TM	Trichloroethene	0.335	0.353	-5.4	88	0.00
45 C	1,2-Dichloropropane	0.364	0.380	-4.4#	86	0.00
46 T	Dibromomethane	0.253	0.267	-5.5	87	0.00
47 T	Bromodichloromethane	0.489	0.537	-9.8	90	0.00
48 T	Methyl methacrylate	0.397	0.433	-9.1	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044948.D
 Acq On : 14 Feb 2025 10:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 22:10:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	85	0.00
50 S	Toluene-d8	1.229	1.231	-0.2	83	0.00
51 T	4-Methyl-2-Pentanone	0.512	0.589	-15.0	89	0.00
52 CM	Toluene	0.872	0.912	-4.6#	85	0.00
53 T	t-1,3-Dichloropropene	0.495	0.526	-6.3	85	0.00
54 T	cis-1,3-Dichloropropene	0.552	0.593	-7.4	86	0.00
55 T	1,1,2-Trichloroethane	0.335	0.359	-7.2	88	0.00
56 T	Ethyl methacrylate	0.542	0.582	-7.4	84	0.00
57 T	1,3-Dichloropropane	0.587	0.630	-7.3	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.289	-8.6	84	0.00
59 T	2-Hexanone	0.369	0.430	-16.5	89	0.00
60 T	Dibromochloromethane	0.359	0.395	-10.0	88	0.00
61 T	1,2-Dibromoethane	0.340	0.361	-6.2	86	0.00
62 S	4-Bromofluorobenzene	0.414	0.440	-6.3	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.315	0.332	-5.4	89	0.00
65 PM	Chlorobenzene	1.074	1.118	-4.1	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.346	0.371	-7.2	88	0.00
67 C	Ethyl Benzene	1.895	1.994	-5.2#	86	0.00
68 T	m/p-Xylenes	0.699	0.739	-5.7	85	0.00
69 T	o-Xylene	0.701	0.733	-4.6	87	0.00
70 T	Styrene	1.130	1.244	-10.1	87	0.00
71 P	Bromoform	0.258	0.294	-14.0	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	4.026	4.087	-1.5	86	0.00
74 T	N-amyl acetate	1.822	1.921	-5.4	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.383	1.368	1.1	86	0.00
76 T	1,2,3-Trichloropropane	1.109	1.135	-2.3	88	0.00
77 T	Bromobenzene	0.913	0.945	-3.5	89	0.00
78 T	n-propylbenzene	4.647	4.842	-4.2	88	0.00
79 T	2-Chlorotoluene	2.856	2.873	-0.6	86	0.00
80 T	1,3,5-Trimethylbenzene	3.267	3.407	-4.3	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.436	-1.2	85	0.00
82 T	4-Chlorotoluene	3.160	3.288	-4.1	88	0.00
83 T	tert-Butylbenzene	3.305	3.425	-3.6	88	0.00
84 T	1,2,4-Trimethylbenzene	3.225	3.455	-7.1	88	0.00
85 T	sec-Butylbenzene	4.093	4.362	-6.6	90	0.00
86 T	p-Isopropyltoluene	3.291	3.553	-8.0	89	0.00
87 T	1,3-Dichlorobenzene	1.678	1.756	-4.6	89	0.00
88 T	1,4-Dichlorobenzene	1.697	1.735	-2.2	88	0.00
89 T	n-Butylbenzene	2.912	3.285	-12.8	90	0.00
90 T	Hexachloroethane	0.616	0.637	-3.4	89	0.00
91 T	1,2-Dichlorobenzene	1.650	1.705	-3.3	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.252	0.271	-7.5	90	0.00
93 T	1,2,4-Trichlorobenzene	0.998	1.054	-5.6	89	0.00
94 T	Hexachlorobutadiene	0.396	0.426	-7.6	95	0.00
95 T	Naphthalene	3.608	3.772	-4.5	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044948.D
Acq On : 14 Feb 2025 10:04
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 14 22:10:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.006	1.065	-5.9	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6