

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044974.D
 Acq On : 14 Feb 2025 20:06
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 22:21:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	74	0.00
2 T	Dichlorodifluoromethane	0.701	0.706	-0.7	74	0.00
3 P	Chloromethane	0.854	0.843	1.3	74	0.00
4 C	Vinyl Chloride	0.827	0.791	4.4#	72	0.00
5 T	Bromomethane	0.247	0.344	-39.3#	102	0.00
6 T	Chloroethane	0.307	0.461	-50.2#	100	0.01
7 T	Trichlorofluoromethane	1.046	1.106	-5.7	79	0.00
8 T	Diethyl Ether	0.379	0.404	-6.6	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.617	2.4	73	0.00
10 T	Methyl Iodide	0.819	0.868	-6.0	76	0.00
11 T	Tert butyl alcohol	0.136	0.173	-27.2#	94	-0.02
12 CM	1,1-Dichloroethene	0.644	0.650	-0.9#	76	0.00
13 T	Acrolein	0.140	0.155	-10.7	81	0.00
14 T	Allyl chloride	1.165	1.226	-5.2	78	0.00
15 T	Acrylonitrile	0.363	0.421	-16.0	82	0.00
16 T	Acetone	0.294	0.359	-22.1	90	0.00
17 T	Carbon Disulfide	1.767	1.674	5.3	70	0.00
18 T	Methyl Acetate	0.928	1.085	-16.9	85	0.00
19 T	Methyl tert-butyl Ether	2.050	2.224	-8.5	80	0.00
20 T	Methylene Chloride	0.721	0.748	-3.7	78	0.00
21 T	trans-1,2-Dichloroethene	0.634	0.651	-2.7	75	0.00
22 T	Diisopropyl ether	2.203	2.433	-10.4	80	0.00
23 T	Vinyl Acetate	1.796	2.035	-13.3	81	0.00
24 P	1,1-Dichloroethane	1.233	1.321	-7.1	79	0.00
25 T	2-Butanone	0.478	0.578	-20.9	84	0.00
26 T	2,2-Dichloropropane	0.978	0.902	7.8	70	0.00
27 T	cis-1,2-Dichloroethene	0.762	0.803	-5.4	78	0.00
28 T	Bromochloromethane	0.593	0.552	6.9	70	0.00
29 T	Tetrahydrofuran	0.317	0.378	-19.2	85	0.00
30 C	Chloroform	1.196	1.296	-8.4#	82	0.00
31 T	Cyclohexane	1.137	1.103	3.0	73	0.00
32 T	1,1,1-Trichloroethane	1.014	1.079	-6.4	79	0.00
33 S	1,2-Dichloroethane-d4	0.732	0.698	4.6	71	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.325	0.301	7.4	69	0.00
36 T	1,1-Dichloropropene	0.469	0.483	-3.0	77	0.00
37 T	Ethyl Acetate	0.522	0.594	-13.8	83	0.00
38 T	Carbon Tetrachloride	0.460	0.487	-5.9	79	0.00
39 T	Methylcyclohexane	0.606	0.598	1.3	71	0.00
40 TM	Benzene	1.465	1.543	-5.3	77	0.00
41 T	Methacrylonitrile	0.279	0.349	-25.1#	85	0.00
42 TM	1,2-Dichloroethane	0.467	0.534	-14.3	83	0.00
43 T	Isopropyl Acetate	0.842	0.973	-15.6	82	0.00
44 TM	Trichloroethene	0.335	0.359	-7.2	79	0.00
45 C	1,2-Dichloropropane	0.364	0.398	-9.3#	80	0.00
46 T	Dibromomethane	0.253	0.283	-11.9	81	0.00
47 T	Bromodichloromethane	0.489	0.560	-14.5	82	0.00
48 T	Methyl methacrylate	0.397	0.479	-20.7	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
 Data File : VX044974.D
 Acq On : 14 Feb 2025 20:06
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 22:21:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.011	-22.2	89	0.00
50 S	Toluene-d8	1.229	1.100	10.5	65	0.00
51 T	4-Methyl-2-Pentanone	0.512	0.633	-23.6	84	0.00
52 CM	Toluene	0.872	0.928	-6.4#	76	0.00
53 T	t-1,3-Dichloropropene	0.495	0.551	-11.3	79	0.00
54 T	cis-1,3-Dichloropropene	0.552	0.604	-9.4	77	0.00
55 T	1,1,2-Trichloroethane	0.335	0.374	-11.6	80	0.00
56 T	Ethyl methacrylate	0.542	0.619	-14.2	79	0.00
57 T	1,3-Dichloropropane	0.587	0.647	-10.2	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.288	-8.3	73	0.00
59 T	2-Hexanone	0.369	0.479	-29.8#	87	0.00
60 T	Dibromochloromethane	0.359	0.410	-14.2	81	0.00
61 T	1,2-Dibromoethane	0.340	0.382	-12.4	80	0.00
62 S	4-Bromofluorobenzene	0.414	0.407	1.7	69	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	75	0.00
64 T	Tetrachloroethene	0.315	0.320	-1.6	77	0.00
65 PM	Chlorobenzene	1.074	1.141	-6.2	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.346	0.376	-8.7	80	0.00
67 C	Ethyl Benzene	1.895	2.011	-6.1#	78	0.00
68 T	m/p-Xylenes	0.699	0.741	-6.0	77	0.00
69 T	o-Xylene	0.701	0.735	-4.9	78	0.00
70 T	Styrene	1.130	1.247	-10.4	78	0.00
71 P	Bromoform	0.258	0.302	-17.1	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	4.026	3.800	5.6	77	0.00
74 T	N-amyl acetate	1.822	1.942	-6.6	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.383	1.364	1.4	82	0.00
76 T	1,2,3-Trichloropropane	1.109	1.149	-3.6	85	0.00
77 T	Bromobenzene	0.913	0.907	0.7	82	0.00
78 T	n-propylbenzene	4.647	4.561	1.9	79	0.00
79 T	2-Chlorotoluene	2.856	2.729	4.4	79	0.00
80 T	1,3,5-Trimethylbenzene	3.267	3.167	3.1	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.389	9.7	73	0.00
82 T	4-Chlorotoluene	3.160	3.107	1.7	80	0.00
83 T	tert-Butylbenzene	3.305	3.218	2.6	80	0.00
84 T	1,2,4-Trimethylbenzene	3.225	3.227	-0.1	79	0.00
85 T	sec-Butylbenzene	4.093	4.055	0.9	80	0.00
86 T	p-Isopropyltoluene	3.291	3.340	-1.5	81	0.00
87 T	1,3-Dichlorobenzene	1.678	1.681	-0.2	82	0.00
88 T	1,4-Dichlorobenzene	1.697	1.686	0.6	82	0.00
89 T	n-Butylbenzene	2.912	3.218	-10.5	85	0.00
90 T	Hexachloroethane	0.616	0.587	4.7	79	0.00
91 T	1,2-Dichlorobenzene	1.650	1.679	-1.8	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.252	0.285	-13.1	91	0.00
93 T	1,2,4-Trichlorobenzene	0.998	1.076	-7.8	87	0.00
94 T	Hexachlorobutadiene	0.396	0.416	-5.1	89	0.00
95 T	Naphthalene	3.608	3.854	-6.8	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021425\
Data File : VX044974.D
Acq On : 14 Feb 2025 20:06
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 14 22:21:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.006	1.055	-4.9	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6