

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021519\
 Data File : VX007538.D
 Acq On : 14 Feb 2019 11:55
 Operator : JC/SP
 Sample : VX0215WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0215WBL01

Quant Time: Feb 15 02:46:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	479807	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	762159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	713813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	331388	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	267764	52.58	ug/l	0.00
Spiked Amount			Recovery	=	105.16%	
35) Dibromofluoromethane	5.50	113	249958	50.40	ug/l	0.00
Spiked Amount			Recovery	=	100.80%	
50) Toluene-d8	8.71	98	908931	49.40	ug/l	0.00
Spiked Amount			Recovery	=	98.80%	
62) 4-Bromofluorobenzene	11.13	95	312624	46.54	ug/l	0.00
Spiked Amount			Recovery	=	93.08%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021519\
Data File : VX007538.D
Acq On : 14 Feb 2019 11:55
Operator : JC/SP
Sample : VX0215WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0215WBL01

Quant Time: Feb 15 02:46:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

