

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021519\
 Data File : VX007542.D
 Acq On : 14 Feb 2019 13:42
 Operator : JC/SP
 Sample : K1458-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-6

Quant Time: Feb 15 04:56:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	479696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	752826	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	742460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	350421	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	266690	52.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.76%	
35) Dibromofluoromethane	5.51	113	248685	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
50) Toluene-d8	8.71	98	938017	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
62) 4-Bromofluorobenzene	11.13	95	330889	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
Target Compounds						
16) Acetone	2.45	43	33178	14.225	ug/l	95
20) Methylene Chloride	2.86	84	81838	16.918	ug/l	95
43) Isopropyl Acetate	6.46	43	28714	2.724	ug/l	99
95) Naphthalene	13.83	128	239009	10.114	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021519\
Data File : VX007542.D
Acq On : 14 Feb 2019 13:42
Operator : JC/SP
Sample : K1458-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-6

Quant Time: Feb 15 04:56:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

