

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021524\
 Data File : VX040253.D
 Acq On : 15 Feb 2024 15:37
 Operator : JC/MD
 Sample : PB159010TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159010TB

Quant Time: Feb 16 00:50:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	129744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130837	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68773	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48766	44.171	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.340%
35) Dibromofluoromethane	5.385	113	42181	46.554	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.100%
50) Toluene-d8	8.647	98	160995	47.501	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.000%
62) 4-Bromofluorobenzene	11.079	95	64973	47.004	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.000%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	12480	27.523	ug/l	90
20) Methylene Chloride	2.794	84	3026	2.030	ug/l	95
43) Isopropyl Acetate	6.336	43	57259	25.972	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021524\
Data File : VX040253.D
Acq On : 15 Feb 2024 15:37
Operator : JC/MD
Sample : PB159010TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159010TB

Quant Time: Feb 16 00:50:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

