

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021524\
 Data File : VX040255.D
 Acq On : 15 Feb 2024 16:22
 Operator : JC/MD
 Sample : PB159010ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159010ZHE#02

Quant Time: Feb 16 00:51:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	128774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	49728	45.066	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	90.140%		
35) Dibromofluoromethane	5.391	113	41552	46.205	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.420%		
50) Toluene-d8	8.647	98	161764	48.087	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.180%		
62) 4-Bromofluorobenzene	11.079	95	63513	46.294	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.580%		
Target Compounds						
16) Acetone	2.380	43	13251	29.239	ug/l	91
20) Methylene Chloride	2.788	84	3351	2.427	ug/l	88
43) Isopropyl Acetate	6.336	43	59293	27.097	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021524\
Data File : VX040255.D
Acq On : 15 Feb 2024 16:22
Operator : JC/MD
Sample : PB159010ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159010ZHE#02

Quant Time: Feb 16 00:51:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

