

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021524\
 Data File : VX040258.D
 Acq On : 15 Feb 2024 17:31
 Operator : JC/MD
 Sample : P1461-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2

Quant Time: Feb 16 00:53:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72755	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	124961	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48376	44.694	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.380%
35) Dibromofluoromethane	5.391	113	40110	45.963	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.920%
50) Toluene-d8	8.647	98	154315	47.273	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.540%
62) 4-Bromofluorobenzene	11.079	95	61770	46.397	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	13404	30.152	ug/l	# 86
20) Methylene Chloride	2.788	84	2711	1.713	ug/l	# 78
43) Isopropyl Acetate	6.342	43	57614	27.133	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021524\
Data File : VX040258.D
Acq On : 15 Feb 2024 17:31
Operator : JC/MD
Sample : P1461-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2

Quant Time: Feb 16 00:53:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

