

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021524\
 Data File : VX040259.D
 Acq On : 15 Feb 2024 17:54
 Operator : JC/MD
 Sample : P1461-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1

Quant Time: Feb 16 00:53:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73005	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	124348	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48090	44.278	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.560%
35) Dibromofluoromethane	5.391	113	40279	46.384	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.760%
50) Toluene-d8	8.647	98	152936	47.081	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.160%
62) 4-Bromofluorobenzene	11.079	95	60946	46.004	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.000%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	12408	27.816	ug/l	95
20) Methylene Chloride	2.788	84	2698	1.686	ug/l	96
43) Isopropyl Acetate	6.336	43	55028	26.043	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021524\
Data File : VX040259.D
Acq On : 15 Feb 2024 17:54
Operator : JC/MD
Sample : P1461-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1

Quant Time: Feb 16 00:53:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

