

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021619\
 Data File : VX007570.D
 Acq On : 15 Feb 2019 16:11
 Operator : JC/SP
 Sample : K1458-07DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-4DL

Quant Time: Feb 16 04:19:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	493279	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	777947	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	720871	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	342121	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	272150	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
35) Dibromofluoromethane	5.50	113	254749	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
50) Toluene-d8	8.71	98	918312	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	11.13	95	324621	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
Target Compounds						
16) Acetone	2.45	43	7928	3.306	ug/l	97
20) Methylene Chloride	2.86	84	164157	33.379	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021619\
Data File : VX007570.D
Acq On : 15 Feb 2019 16:11
Operator : JC/SP
Sample : K1458-07DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-4DL

Quant Time: Feb 16 04:19:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

