


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021619\
Data File : VX007559.D
Acq On : 15 Feb 2019 09:13
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Title : SW846 8260
Last Update : Sat Feb 02 01:32:36 2019



AutoFind: Scans 1623, 1624, 1625; Background Corrected with Scan 1617

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	6173	PASS
75	95	30	60	47.2	18272	PASS
95	95	100	100	100.0	38702	PASS
96	95	5	9	6.5	2512	PASS
173	174	0.00	2	1.8	602	PASS
174	95	50	100	87.9	34029	PASS
175	174	5	9	8.2	2788	PASS
176	174	95	101	96.6	32858	PASS
177	176	5	9	6.4	2103	PASS