

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027047.D
 Acq On : 16 Feb 2022 13:48
 Operator : JC/MD
 Sample : VX0216WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0216WBS02

Quant Time: Feb 17 07:09:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	86850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	138568	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61529	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	53492	51.309	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.620%	
35) Dibromofluoromethane	5.385	113	42937	52.010	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.020%	
50) Toluene-d8	8.647	98	161727	53.368	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.740%	
62) 4-Bromofluorobenzene	11.079	95	63394	59.419	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 118.840%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	17714	18.758	ug/l	99
3) Chloromethane	1.294	50	16628	18.927	ug/l	99
4) Vinyl Chloride	1.374	62	16631	18.084	ug/l	98
5) Bromomethane	1.611	94	10509	26.929	ug/l	89
6) Chloroethane	1.691	64	10101	18.272	ug/l	97
7) Trichlorofluoromethane	1.892	101	26023	19.228	ug/l	98
8) Diethyl Ether	2.136	74	9037	18.026	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.331	101	17651	19.897	ug/l	99
10) Methyl Iodide	2.453	142	19650	17.504	ug/l	98
11) Tert butyl alcohol	2.959	59	16276	86.418	ug/l	98
12) 1,1-Dichloroethene	2.325	96	16533	18.601	ug/l	93
13) Acrolein	2.239	56	12573	72.238	ug/l	100
14) Allyl chloride	2.666	41	29428	18.691	ug/l	96
15) Acrylonitrile	3.062	53	50793	94.112	ug/l	100
16) Acetone	2.380	43	51467	112.399	ug/l	99
17) Carbon Disulfide	2.514	76	40783	16.208	ug/l	99
18) Methyl Acetate	2.703	43	23131	17.557	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	57171	18.172	ug/l	99
20) Methylene Chloride	2.788	84	18686	18.834	ug/l	95
21) trans-1,2-Dichloroethene	3.093	96	17614	18.832	ug/l	98
22) Diisopropyl ether	3.757	45	60811	19.185	ug/l	91
23) Vinyl Acetate	3.721	43	248796	93.354	ug/l	98
24) 1,1-Dichloroethane	3.611	63	32878	19.168	ug/l	99
25) 2-Butanone	4.550	43	72534	98.319	ug/l	98
26) 2,2-Dichloropropane	4.483	77	23079	19.915	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	20932	19.492	ug/l	95
28) Bromochloromethane	4.897	49	13989	19.675	ug/l	96
29) Tetrahydrofuran	5.001	42	46881	93.474	ug/l	99
30) Chloroform	5.099	83	33744	18.935	ug/l	96
31) Cyclohexane	5.471	56	30486	18.403	ug/l	98
32) 1,1,1-Trichloroethane	5.385	97	28930	18.413	ug/l	100
36) 1,1-Dichloropropene	5.696	75	25101	19.273	ug/l	98
37) Ethyl Acetate	4.715	43	27151	18.344	ug/l	100
38) Carbon Tetrachloride	5.678	117	26178	19.168	ug/l	96
39) Methylcyclohexane	7.379	83	31226	19.551	ug/l	95
40) Benzene	6.037	78	74309	19.577	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027047.D
 Acq On : 16 Feb 2022 13:48
 Operator : JC/MD
 Sample : VX0216WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0216WBS02

Quant Time: Feb 17 07:09:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	14882	18.552	ug/l	95
42) 1,2-Dichloroethane	6.086	62	27248	20.372	ug/l	98
43) Isopropyl Acetate	6.336	43	42599	18.706	ug/l	98
44) Trichloroethene	7.129	130	21453	18.823	ug/l	94
45) 1,2-Dichloropropane	7.434	63	18491	19.254	ug/l	98
46) Dibromomethane	7.580	93	13348	20.087	ug/l	97
47) Bromodichloromethane	7.824	83	25805	20.278	ug/l	98
48) Methyl methacrylate	7.690	41	20867	18.493	ug/l	96
49) 1,4-Dioxane	7.665	88	7615	391.632	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	140353	101.704	ug/l	98
52) Toluene	8.720	92	46494	19.347	ug/l	96
53) t-1,3-Dichloropropene	8.982	75	24682	18.783	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	28092	19.991	ug/l	94
55) 1,1,2-Trichloroethane	9.153	97	19507	21.441	ug/l	97
56) Ethyl methacrylate	9.116	69	28696	19.898	ug/l	97
57) 1,3-Dichloropropane	9.311	76	32625	20.883	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	61307	82.824	ug/l	98
59) 2-Hexanone	9.433	43	107850	106.846	ug/l	98
60) Dibromochloromethane	9.519	129	19981	21.345	ug/l	99
61) 1,2-Dibromoethane	9.610	107	19843	20.643	ug/l	100
64) Tetrachloroethene	9.275	164	22777	17.831	ug/l	94
65) Chlorobenzene	10.079	112	51785	19.688	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	18802	19.432	ug/l	98
67) Ethyl Benzene	10.195	91	92490	19.868	ug/l	100
68) m/p-Xylenes	10.305	106	72010	40.204	ug/l	97
69) o-Xylene	10.640	106	35214	20.064	ug/l	95
70) Styrene	10.659	104	58706	20.624	ug/l	99
71) Bromoform	10.799	173	13693	19.453	ug/l #	99
73) Isopropylbenzene	10.963	105	92824	19.153	ug/l	99
74) N-amyl acetate	10.842	43	35497	18.331	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	26411	20.081	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	31212	23.173	ug/l	93
77) Bromobenzene	11.201	156	22418	19.460	ug/l	95
78) n-propylbenzene	11.305	91	107246	19.961	ug/l	99
79) 2-Chlorotoluene	11.366	91	65116	19.615	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	78410	19.658	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	6938	18.406	ug/l	96
82) 4-Chlorotoluene	11.457	91	73889	19.840	ug/l	96
83) tert-Butylbenzene	11.713	119	74190	19.515	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	78801	19.586	ug/l	99
85) sec-Butylbenzene	11.890	105	97122	20.154	ug/l	100
86) p-Isopropyltoluene	12.012	119	82115	20.114	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	42787	20.269	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	42416	19.924	ug/l	98
89) n-Butylbenzene	12.335	91	68319	20.453	ug/l	99
90) Hexachloroethane	12.542	117	12346	20.924	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	41217	20.236	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	5649	17.466	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	24802	19.294	ug/l	98
94) Hexachlorobutadiene	13.725	225	10872	19.995	ug/l	97
95) Naphthalene	13.774	128	83108	18.457	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	25275	19.418	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027047.D
Acq On : 16 Feb 2022 13:48
Operator : JC/MD
Sample : VX0216WBS02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Quant Time: Feb 17 07:09:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027047.D
Acq On : 16 Feb 2022 13:48
Operator : JC/MD
Sample : VX0216WBS02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Quant Time: Feb 17 07:09:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion:168 Resp: 86850
Ion Ratio Lower Upper
168 100
99 57.9 50.3 75.5






#2
Dichlorodifluoromethane
Concen: 18.758 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 85 Resp: 17714
Ion Ratio Lower Upper
85 100
87 33.3 16.4 49.1



Abundance Scan 34 (1.294 min): VX026805.D\data.ms (-29)



#3

Chloromethane

Concen: 18.927 ug/l

RT: 1.294 min Scan# 34

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :


VX0216WBS02

Tgt Ion: 50 Resp: 16628




Ion Ratio Lower Upper

50 100

52 32.9 26.1 39.1



Abundance Scan 48 (1.380 min): VX026805.D\data.ms (-42)



#4

Vinyl Chloride

Concen: 18.084 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.006 min

Lab File: VX027047.D


Acq: 16 Feb 2022 13:48

Tgt Ion: 62 Resp: 16631

Ion Ratio Lower Upper

62 100

64 31.7 26.2 39.2









#5
Bromomethane
Concen: 26.929 ug/l
RT: 1.611 min Scan# 80
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 94 Resp: 10509
Ion Ratio Lower Upper
94 100
96 84.7 75.9 113.9




#6
Chloroethane
Concen: 18.272 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 64 Resp: 10101
Ion Ratio Lower Upper
64 100
66 33.2 25.4 38.2



Abundance Scan 132 (1.892 min): VX026805.D\data.ms (-12



#7

Trichlorofluoromethane

Concen: 19.228 ug/l

RT: 1.892 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :

VX0216WBS02


Tgt Ion:101 Resp: 26023

Ion Ratio Lower Upper


101 100

103 62.8 51.3 76.9


Abundance Scan 132 (1.892 min): VX027047.D\data.ms




Abundance



Abundance Scan 132 (1.892 min): VX027047.D\data.ms (-83



Abundance Scan 172 (2.136 min): VX026805.D\data.ms (-16



#8

Diethyl Ether

Concen: 18.026 ug/l

RT: 2.136 min Scan# 172

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48


Tgt Ion: 74 Resp: 9037

Ion Ratio Lower Upper


74 100

45 97.8 50.7 152.1


Abundance Scan 172 (2.136 min): VX027047.D\data.ms




Abundance



Abundance Scan 172 (2.136 min): VX027047.D\data.ms (-12









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.897 ug/l
RT: 2.331 min Scan# 205
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:	101	Resp:	17651
Ion	Ratio	Lower	Upper
101	100		
85	43.2	35.3	52.9
151	78.0	61.4	92.2



#10
Methyl Iodide
Concen: 17.504 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:	142	Resp:	19650
Ion	Ratio	Lower	Upper
142	100		
127	45.6	38.1	57.1
141	14.0	11.2	16.8





#11

Tert butyl alcohol

Concen: 86.418 ug/l

RT: 2.959 min Scan# 30

Delta R.T. -0.042 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :





VX0216WBS02

Tgt Ion: 59 Resp: 16276

Ion Ratio Lower Upper

59 100

57 9.7 7.2 10.8



#12

1,1-Dichloroethene

Concen: 18.601 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48




Tgt Ion: 96 Resp: 16533


Ion Ratio Lower Upper

96 100

61 161.7 138.5 207.7

98 67.3 51.8 77.6





#13

Acrolein

Concen: 72.238 ug/l

RT: 2.239 min Scan# 189

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :





VX0216WBS02

Tgt Ion: 56 Resp: 12573

Ion Ratio Lower Upper

56 100

55 71.2 56.9 85.3



#14

Allyl chloride

Concen: 18.691 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48




Tgt Ion: 41 Resp: 29428

Ion Ratio Lower Upper

41 100

39 64.2 54.2 81.4

76 32.5 25.2 37.8









#15
Acrylonitrile
Concen: 94.112 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:	53	Resp:	50793
Ion	Ratio	Lower	Upper
53	100		
52	83.0	66.6	100.0
51	36.3	29.5	44.3



#16
Acetone
Concen: 112.399 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:	43	Resp:	51467
Ion	Ratio	Lower	Upper
43	100		
58	29.4	23.0	34.6









#17
Carbon Disulfide
Concen: 16.208 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion: 76 Resp: 40783
Ion Ratio Lower Upper
76 100
78 9.0 7.5 11.3



#18
Methyl Acetate
Concen: 17.557 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 43 Resp: 23131
Ion Ratio Lower Upper
43 100
74 23.7 18.3 27.5








#19
Methyl tert-butyl Ether
Concen: 18.172 ug/l
RT: 3.111 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 73 Resp: 57171
Ion Ratio Lower Upper
73 100
57 22.5 17.4 26.2



#20
Methylene Chloride
Concen: 18.834 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 84 Resp: 18686
Ion Ratio Lower Upper
84 100
49 120.1 101.5 152.3
51 34.7 31.2 46.8
86 63.4 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 18.832 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 96 Resp: 17614
Ion Ratio Lower Upper
96 100
61 139.5 112.9 169.3
98 64.5 49.5 74.3



#22
Diisopropyl ether
Concen: 19.185 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 45 Resp: 60811
Ion Ratio Lower Upper
45 100
43 87.3 60.6 91.0
87 28.3 22.0 33.0
59 11.8 9.0 13.4









#23
 Vinyl Acetate
 Concen: 93.354 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX027047.D
 Acq: 16 Feb 2022 13:48




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0216WBS02


Tgt Ion: 43 Resp: 248796
 Ion Ratio Lower Upper
 43 100
 86 10.9 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 19.168 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX027047.D
 Acq: 16 Feb 2022 13:48

Tgt Ion: 63 Resp: 32878
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.7 11.1
 100 4.9 2.4 7.2









#25
2-Butanone
Concen: 98.319 ug/l
RT: 4.550 min Scan# 570
Delta R.T. -0.012 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion: 43 Resp: 72534
Ion Ratio Lower Upper
43 100
72 24.3 20.4 30.6



#26
2,2-Dichloropropane
Concen: 19.915 ug/l
RT: 4.483 min Scan# 557
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 77 Resp: 23079
Ion Ratio Lower Upper
77 100
97 24.0 11.0 33.0








#27
 cis-1,2-Dichloroethene
 Concen: 19.492 ug/l
 RT: 4.489 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX027047.D
 Acq: 16 Feb 2022 13:48




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0216WBS02


Tgt Ion: 96 Resp: 20932
 Ion Ratio Lower Upper
 96 100
 61 140.4 0.0 296.6
 98 63.6 0.0 128.8



#28
 Bromochloromethane
 Concen: 19.675 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. -0.006 min
 Lab File: VX027047.D
 Acq: 16 Feb 2022 13:48

Tgt Ion: 49 Resp: 13989
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 4.2
 130 81.3 62.4 93.6





#29

Tetrahydrofuran

Concen: 93.474 ug/l

RT: 5.001 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :

VX0216WBS02





Tgt Ion: 42 Resp: 46881

Ion Ratio Lower Upper

42 100

72 44.0 34.6 52.0

71 40.1 31.8 47.8



#30

Chloroform

Concen: 18.935 ug/l

RT: 5.099 min Scan# 658

Delta R.T. 0.000 min

Lab File: VX027047.D




Acq: 16 Feb 2022 13:48


Tgt Ion: 83 Resp: 33744

Ion Ratio Lower Upper

83 100

85 68.0 51.6 77.4









#31
Cyclohexane
Concen: 18.403 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion: 56 Resp: 30486
Ion Ratio Lower Upper
56 100
69 30.9 25.8 38.6
84 88.0 68.9 103.3



#32
1,1,1-Trichloroethane
Concen: 18.413 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 97 Resp: 28930
Ion Ratio Lower Upper
97 100
99 64.2 51.3 76.9
61 44.3 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 51.309 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :





VX0216WBS02

Tgt Ion: 65 Resp: 53492

Ion Ratio Lower Upper

65 100

67 51.1 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48




Tgt Ion: 114 Resp: 138568


Ion Ratio Lower Upper

114 100

63 20.1 0.0 42.8

88 16.1 0.0 33.0









#35
Dibromofluoromethane
Concen: 52.010 ug/l
RT: 5.385 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:113 Resp: 42937
Ion Ratio Lower Upper
113 100
111 103.4 81.2 121.8
192 19.5 15.0 22.4



#36
1,1-Dichloropropene
Concen: 19.273 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 75 Resp: 25101
Ion Ratio Lower Upper
75 100
110 35.5 17.2 51.6
77 31.5 24.6 36.8





#37

Ethyl Acetate

Concen: 18.344 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :

VX0216WBS02





Tgt Ion: 43 Resp: 27151

Ion Ratio Lower Upper

43 100

61 13.3 10.6 16.0

70 10.6 8.5 12.7



#38

Carbon Tetrachloride

Concen: 19.168 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48




Tgt Ion: 117 Resp: 26178


Ion Ratio Lower Upper

117 100

119 98.4 74.7 112.1

121 30.8 25.4 38.0





#39

Methylcyclohexane

Concen: 19.551 ug/l

RT: 7.379 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :

VX0216WBS02





Tgt Ion: 83 Resp: 31226

Ion Ratio Lower Upper

83 100

55 74.5 63.6 95.4

98 50.4 38.4 57.6



#40

Benzene

Concen: 19.577 ug/l

RT: 6.037 min Scan# 812

Delta R.T. -0.006 min

Lab File: VX027047.D




Acq: 16 Feb 2022 13:48


Tgt Ion: 78 Resp: 74309

Ion Ratio Lower Upper

78 100

77 21.2 18.9 28.3









#41
Methacrylonitrile
Concen: 18.552 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 41 Resp: 14882




Ion	Ratio	Lower	Upper
41	100		
39	56.3	45.4	68.0
67	74.3	53.2	79.8
52	28.9	23.0	34.6




#42
1,2-Dichloroethane
Concen: 20.372 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 62 Resp: 27248

Ion	Ratio	Lower	Upper
62	100		
98	9.7	0.0	17.8









#43
Isopropyl Acetate
Concen: 18.706 ug/l
RT: 6.336 min Scan# 861
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion: 43 Resp: 42599
Ion Ratio Lower Upper
43 100
61 20.4 15.4 23.0
87 13.2 10.2 15.2



#44
Trichloroethene
Concen: 18.823 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:130 Resp: 21453
Ion Ratio Lower Upper
130 100
95 95.7 0.0 202.6





#45

1,2-Dichloropropane

Concen: 19.254 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :




VX0216WBS02

Tgt Ion: 63 Resp: 18491

Ion Ratio Lower Upper

63 100

65 31.0 23.8 35.6



#46

Dibromomethane

Concen: 20.087 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48




Tgt Ion: 93 Resp: 13348


Ion Ratio Lower Upper

93 100

95 80.1 66.1 99.1

174 100.9 78.7 118.1









#47
Bromodichloromethane
Concen: 20.278 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 83	Resp: 25805
Ion Ratio	Lower Upper
83 100	
85 62.4	51.4 77.0
127 8.3	6.9 10.3



#48
Methyl methacrylate
Concen: 18.493 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 41	Resp: 20867
Ion Ratio	Lower Upper
41 100	
69 83.4	62.2 93.4
39 55.8	45.4 68.0









#49
1,4-Dioxane
Concen: 391.632 ug/l
RT: 7.665 min Scan# 1081
Delta R.T. -0.012 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 88 Resp: 7615
Ion Ratio Lower Upper
88 100
43 33.8 29.9 44.9
58 70.7 57.8 86.6



#50
Toluene-d8
Concen: 53.368 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 98 Resp: 161727
Ion Ratio Lower Upper
98 100
100 66.3 53.2 79.8









#51
4-Methyl-2-Pentanone
Concen: 101.704 ug/l
RT: 8.574 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion: 43 Resp: 140353
Ion Ratio Lower Upper
43 100
58 39.2 30.5 45.7



#52
Toluene
Concen: 19.347 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 92 Resp: 46494
Ion Ratio Lower Upper
92 100
91 174.6 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 18.783 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :





VX0216WBS02

Tgt Ion: 75 Resp: 24682

Ion Ratio Lower Upper

75 100

77 31.0 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 19.991 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48




Tgt Ion: 75 Resp: 28092


Ion Ratio Lower Upper

75 100

77 30.6 26.2 39.2

39 49.2 43.5 65.3









#55
1,1,2-Trichloroethane
Concen: 21.441 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 97 Resp: 19507




Ion	Ratio	Lower	Upper
97	100		
83	81.3	68.4	102.6
85	52.6	44.0	66.0
99	61.5	49.8	74.6




#56
Ethyl methacrylate
Concen: 19.898 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 69 Resp: 28696

Ion	Ratio	Lower	Upper
69	100		
41	68.5	57.2	85.8
39	36.9	30.9	46.3









#57
1,3-Dichloropropane
Concen: 20.883 ug/l
RT: 9.311 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 76 Resp: 32625
Ion Ratio Lower Upper
76 100
78 31.7 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 82.824 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 63 Resp: 61307
Ion Ratio Lower Upper
63 100
106 29.9 23.0 34.6





#59

2-Hexanone

Concen: 106.846 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :





VX0216WBS02

Tgt Ion: 43 Resp: 107850

Ion Ratio Lower Upper

43 100

58 54.2 26.4 79.2



#60

Dibromochloromethane

Concen: 21.345 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.006 min

Lab File: VX027047.D




Acq: 16 Feb 2022 13:48


Tgt Ion: 129 Resp: 19981

Ion Ratio Lower Upper

129 100

127 76.0 38.5 115.5









#61
1,2-Dibromoethane
Concen: 20.643 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:107 Resp: 19843
Ion Ratio Lower Upper
107 100
109 94.6 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 59.419 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 95 Resp: 63394
Ion Ratio Lower Upper
95 100
174 76.2 0.0 148.6
176 73.1 0.0 141.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:117 Resp: 127184
Ion Ratio Lower Upper
117 100
82 55.2 45.2 67.8
119 32.5 26.2 39.2



#64
Tetrachloroethene
Concen: 17.831 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:164 Resp: 22777
Ion Ratio Lower Upper
164 100
166 123.1 101.8 152.6
129 93.2 83.3 124.9
131 92.0 78.0 117.0









#65
Chlorobenzene
Concen: 19.688 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:112 Resp: 51785
Ion Ratio Lower Upper
112 100
114 31.2 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 19.432 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:131 Resp: 18802
Ion Ratio Lower Upper
131 100
133 94.7 47.9 143.8
119 65.3 31.7 95.1









#67
Ethyl Benzene
Concen: 19.868 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion: 91 Resp: 92490
Ion Ratio Lower Upper
91 100
106 31.1 24.7 37.1



#68
m/p-Xylenes
Concen: 40.204 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 106 Resp: 72010
Ion Ratio Lower Upper
106 100
91 201.4 165.3 247.9









#69
o-Xylene
Concen: 20.064 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:106 Resp: 35214
Ion Ratio Lower Upper
106 100
91 211.9 109.7 329.0



#70
Styrene
Concen: 20.624 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:104 Resp: 58706
Ion Ratio Lower Upper
104 100
78 51.1 42.5 63.7
103 54.0 43.3 64.9









#71
Bromoform
Concen: 19.453 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:173 Resp: 13693
Ion Ratio Lower Upper
173 100
175 48.1 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:152 Resp: 61529
Ion Ratio Lower Upper
152 100
115 71.9 44.2 132.6
150 166.3 0.0 348.2









#73
Isopropylbenzene
Concen: 19.153 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion:105 Resp: 92824
Ion Ratio Lower Upper
105 100
120 26.6 13.1 39.1



#74
N-ethyl acetate
Concen: 18.331 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 43 Resp: 35497
Ion Ratio Lower Upper
43 100
70 42.6 32.9 49.3
55 24.3 18.3 27.5
61 23.9 18.9 28.3









#75
1,1,2,2-Tetrachloroethane
Concen: 20.081 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:	83	Resp:	26411
Ion	Ratio	Lower	Upper
83	100		
131	10.8	5.3	15.9
85	64.9	32.0	96.2



#76
1,2,3-Trichloropropane
Concen: 23.173 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:	75	Resp:	31212
Ion	Ratio	Lower	Upper
75	100		
77	32.3	18.3	54.8









#77
Bromobenzene
Concen: 19.460 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:156 Resp: 22418
Ion Ratio Lower Upper
156 100
77 159.4 84.5 253.4
158 96.9 48.1 144.4



#78
n-propylbenzene
Concen: 19.961 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 91 Resp: 107246
Ion Ratio Lower Upper
91 100
120 23.6 11.6 34.7









#79
2-Chlorotoluene
Concen: 19.615 ug/l
RT: 11.366 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion: 91 Resp: 65116
Ion Ratio Lower Upper
91 100
126 33.6 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 19.658 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 105 Resp: 78410
Ion Ratio Lower Upper
105 100
120 48.9 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 18.406 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion: 75 Resp: 6938
Ion Ratio Lower Upper
75 100
53 122.9 94.0 141.0
89 48.2 37.0 55.4



#82
4-Chlorotoluene
Concen: 19.840 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 91 Resp: 73889
Ion Ratio Lower Upper
91 100
126 30.2 14.1 42.4









#83
tert-Butylbenzene
Concen: 19.515 ug/l
RT: 11.713 min Scan# 1190
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion	Resp	Lower	Upper
119	100		
91	61.2	31.4	94.0
134	24.0	11.8	35.3



#84
1,2,4-Trimethylbenzene
Concen: 19.586 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion	Resp	Lower	Upper
105	100		
120	44.4	21.9	65.5








#85
sec-Butylbenzene
Concen: 20.154 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:105 Resp: 97122
Ion Ratio Lower Upper
105 100
134 20.2 10.1 30.1



#86
p-Isopropyltoluene
Concen: 20.114 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:119 Resp: 82115
Ion Ratio Lower Upper
119 100
134 25.3 13.0 39.0
91 24.1 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 20.269 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48

Instrument :

MSVOA_X

ClientSampleId :

VX0216WBS02





Tgt Ion:146 Resp: 42787

Ion Ratio Lower Upper

146 100

111 40.1 21.3 63.7

148 62.2 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 19.924 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX027047.D

Acq: 16 Feb 2022 13:48




Tgt Ion:146 Resp: 42416


Ion Ratio Lower Upper

146 100

111 40.7 20.2 60.6

148 64.8 31.4 94.0









#89
n-Butylbenzene
Concen: 20.453 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion: 91 Resp: 68319
Ion Ratio Lower Upper
91 100
92 54.8 27.2 81.6
134 26.8 13.1 39.1



#90
Hexachloroethane
Concen: 20.924 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 117 Resp: 12346
Ion Ratio Lower Upper
117 100
201 73.8 35.6 106.8









#91
1,2-Dichlorobenzene
Concen: 20.236 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion:146 Resp: 41217
Ion Ratio Lower Upper
146 100
111 41.1 21.8 65.3
148 63.6 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.466 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion: 75 Resp: 5649
Ion Ratio Lower Upper
75 100
155 86.0 40.0 119.9
157 107.8 51.6 154.8









#93
1,2,4-Trichlorobenzene
Concen: 19.294 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48




Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02


Tgt Ion:180 Resp: 24802
Ion Ratio Lower Upper
180 100
182 95.7 46.7 140.0
145 32.7 16.3 48.8



#94
Hexachlorobutadiene
Concen: 19.995 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:225 Resp: 10872
Ion Ratio Lower Upper
225 100
223 62.2 32.0 96.0
227 62.0 32.3 96.9








#95
Naphthalene
Concen: 18.457 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.006 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Instrument :
MSVOA_X
ClientSampleId :
VX0216WBS02

Tgt Ion:128	Resp:	83108
Ion	Ratio	Lower Upper
128	100	
127	12.5	10.3 15.5
129	10.8	8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 19.418 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX027047.D
Acq: 16 Feb 2022 13:48

Tgt Ion:180	Resp:	25275
Ion	Ratio	Lower Upper
180	100	
182	92.7	47.2 141.6
145	33.0	17.5 52.5

