

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027034.D
 Acq On : 15 Feb 2022 16:53
 Operator : JC/MD
 Sample : VX0216MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0216MBL01

Quant Time: Feb 16 04:06:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76266	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68210	52.405	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.820%
35) Dibromofluoromethane	5.385	113	55772	51.738	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.480%
50) Toluene-d8	8.647	98	208150	52.604	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.200%
62) 4-Bromofluorobenzene	11.079	95	79665	57.185	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027034.D
Acq On : 15 Feb 2022 16:53
Operator : JC/MD
Sample : VX0216MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0216MBL01

Quant Time: Feb 16 04:06:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

