

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027045.D
 Acq On : 16 Feb 2022 12:56
 Operator : JC/MD
 Sample : VX0216MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0216MBL02

Quant Time: Feb 17 07:09:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111858	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	186455	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79579	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70126	52.226	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.460%
35) Dibromofluoromethane	5.385	113	58654	52.801	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.600%
50) Toluene-d8	8.647	98	216210	53.023	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.040%
62) 4-Bromofluorobenzene	11.079	95	82667	57.583	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027045.D
Acq On : 16 Feb 2022 12:56
Operator : JC/MD
Sample : VX0216MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0216MBL02

Quant Time: Feb 17 07:09:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

