

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027046.D
 Acq On : 16 Feb 2022 13:19
 Operator : JC/MD
 Sample : VX0216WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0216WBL02

Quant Time: Feb 17 07:09:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	111175	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	184952	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169533	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78252	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70532	52.851	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.700%
35) Dibromofluoromethane	5.391	113	58852	53.409	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.820%
50) Toluene-d8	8.653	98	215089	53.177	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.360%
62) 4-Bromofluorobenzene	11.079	95	83370	58.545	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027046.D
Acq On : 16 Feb 2022 13:19
Operator : JC/MD
Sample : VX0216WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0216WBL02

Quant Time: Feb 17 07:09:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

