

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027052.D
 Acq On : 16 Feb 2022 16:16
 Operator : JC/MD
 Sample : N1575-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COP-2R

Quant Time: Feb 17 07:13:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	182509	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68643	52.071	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.140%	
35) Dibromofluoromethane	5.391	113	56451	51.916	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.840%	
50) Toluene-d8	8.653	98	214177	53.660	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.320%	
62) 4-Bromofluorobenzene	11.079	95	83613	59.501	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 119.000%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11583	20.005	ug/l	95
18) Methyl Acetate	2.709	43	527	0.316	ug/l #	81
25) 2-Butanone	4.574	43	1346	1.443	ug/l	96
31) Cyclohexane	5.471	56	1080	0.516	ug/l	95
37) Ethyl Acetate	4.721	43	20270	10.398	ug/l	98
39) Methylcyclohexane	7.379	83	620	0.295	ug/l	91
40) Benzene	6.050	78	5272	1.055	ug/l	97
52) Toluene	8.720	92	22509	7.111	ug/l	99
67) Ethyl Benzene	10.195	91	11087	1.779	ug/l	99
68) m/p-Xylenes	10.299	106	18630	7.770	ug/l	95
69) o-Xylene	10.640	106	15968	6.796	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	13936	2.728	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	14457	2.806	ug/l	95
95) Naphthalene	13.774	128	9099	1.578	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027052.D
Acq On : 16 Feb 2022 16:16
Operator : JC/MD
Sample : N1575-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COP-2R

Quant Time: Feb 17 07:13:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

