

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027056.D
 Acq On : 16 Feb 2022 17:48
 Operator : JC/MD
 Sample : N1575-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-9R

Quant Time: Feb 17 07:15:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	111094	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	182247	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79841	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68017	51.004	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.000%	
35) Dibromofluoromethane	5.391	113	56726	52.244	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.480%	
50) Toluene-d8	8.653	98	217224	54.502	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.000%	
62) 4-Bromofluorobenzene	11.079	95	82941	59.108	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 118.220%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5808	9.916	ug/l	98
25) 2-Butanone	4.562	43	1317	1.396	ug/l	98
31) Cyclohexane	5.477	56	83814	39.554	ug/l	95
37) Ethyl Acetate	4.715	43	10655	5.473	ug/l	98
39) Methylcyclohexane	7.385	83	48895	23.276	ug/l	97
40) Benzene	6.044	78	16932	3.392	ug/l	99
52) Toluene	8.720	92	51551	16.310	ug/l	100
67) Ethyl Benzene	10.195	91	1171485	184.935	ug/l	98
68) m/p-Xylenes	10.305	106	879726	360.959	ug/l	95
69) o-Xylene	10.646	106	191051	80.000	ug/l	97
73) Isopropylbenzene	10.963	105	168577	26.806	ug/l	99
78) n-propylbenzene	11.305	91	540588	77.538	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	395138	76.342	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	1676670	321.155	ug/l	96
86) p-Isopropyltoluene	12.012	119	12909	2.437	ug/l #	73
95) Naphthalene	13.774	128	251740	43.086	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027056.D
Acq On : 16 Feb 2022 17:48
Operator : JC/MD
Sample : N1575-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-9R

Quant Time: Feb 17 07:15:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

