

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027060.D
 Acq On : 16 Feb 2022 19:21
 Operator : JC/MD
 Sample : N1575-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-02-16

Quant Time: Feb 17 07:17:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	111047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	186949	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69611	52.221	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.440%	
35) Dibromofluoromethane	5.391	113	57479	51.606	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.220%	
50) Toluene-d8	8.653	98	221053	54.068	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.140%	
62) 4-Bromofluorobenzene	11.079	95	83345	57.902	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.800%	
Target Compounds						
						Qvalue
68) m/p-Xylenes	10.305	106	1360	0.552	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	1299	0.259	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027060.D
Acq On : 16 Feb 2022 19:21
Operator : JC/MD
Sample : N1575-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-02-16

Quant Time: Feb 17 07:17:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

