

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027064.D
 Acq On : 16 Feb 2022 20:53
 Operator : JC/MD
 Sample : N1577-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-1741-1745

Quant Time: Feb 17 07:19:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107180	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176506	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75199	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66075	51.357	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.720%			
35) Dibromofluoromethane	5.391	113	55201	52.493	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.980%			
50) Toluene-d8	8.653	98	203240	52.652	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.300%			
62) 4-Bromofluorobenzene	11.079	95	79504	58.502	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 117.000%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	22946	40.607	ug/l	96
18) Methyl Acetate	2.709	43	2003	1.232	ug/l	95
25) 2-Butanone	4.568	43	12190	13.389	ug/l	97
31) Cyclohexane	5.477	56	2137	1.045	ug/l #	74
40) Benzene	6.044	78	211012	43.643	ug/l	100
52) Toluene	8.720	92	83528	27.287	ug/l	100
68) m/p-Xylenes	10.299	106	9555	4.185	ug/l	97
69) o-Xylene	10.647	106	3152	1.409	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027064.D
Acq On : 16 Feb 2022 20:53
Operator : JC/MD
Sample : N1577-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-1741-1745

Quant Time: Feb 17 07:19:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

