

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027042.D
 Acq On : 15 Feb 2022 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 16 04:09:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.544	0.516	5.1	81	0.00
3 P	Chloromethane	0.506	0.506	0.0	88	0.00
4 C	Vinyl Chloride	0.529	0.486	8.1#	79	0.00
5 T	Bromomethane	0.225	0.239	-6.2	98	0.00
6 T	Chloroethane	0.281	0.293	-4.3	81	0.00
7 T	Trichlorofluoromethane	0.779	0.746	4.2	81	0.00
8 T	Diethyl Ether	0.289	0.269	6.9	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.511	0.502	1.8	85	0.00
10 T	Methyl Iodide	0.656	0.670	-2.1	84	0.00
11 T	Tert butyl alcohol	0.108	0.085	21.3	70	-0.03
12 CM	1,1-Dichloroethene	0.512	0.482	5.9#	81	0.00
13 T	Acrolein	0.100	0.091	9.0	79	0.00
14 T	Allyl chloride	0.906	0.868	4.2	82	0.00
15 T	Acrylonitrile	0.311	0.289	7.1	80	0.00
16 T	Acetone	0.264	0.237	10.2	81	0.00
17 T	Carbon Disulfide	1.449	1.182	18.4	71	0.00
18 T	Methyl Acetate	0.758	0.654	13.7	78	0.00
19 T	Methyl tert-butyl Ether	1.811	1.701	6.1	82	0.00
20 T	Methylene Chloride	0.571	0.548	4.0	86	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.516	4.1	82	0.00
22 T	Diisopropyl ether	1.825	1.749	4.2	83	0.00
23 T	Vinyl Acetate	1.534	1.491	2.8	82	0.00
24 P	1,1-Dichloroethane	0.987	0.964	2.3	84	0.00
25 T	2-Butanone	0.425	0.393	7.5	80	0.00
26 T	2,2-Dichloropropane	0.667	0.692	-3.7	87	0.00
27 T	cis-1,2-Dichloroethene	0.618	0.627	-1.5	86	0.00
28 T	Bromochloromethane	0.409	0.385	5.9	81	0.00
29 T	Tetrahydrofuran	0.289	0.260	10.0	78	0.00
30 C	Chloroform	1.026	0.987	3.8#	85	0.00
31 T	Cyclohexane	0.954	0.887	7.0	80	0.00
32 T	1,1,1-Trichloroethane	0.905	0.877	3.1	83	0.00
33 S	1,2-Dichloroethane-d4	0.600	0.599	0.2	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.298	0.302	-1.3	92	0.00
36 T	1,1-Dichloropropene	0.470	0.448	4.7	82	0.00
37 T	Ethyl Acetate	0.534	0.478	10.5	77	0.00
38 T	Carbon Tetrachloride	0.493	0.471	4.5	82	0.00
39 T	Methylcyclohexane	0.576	0.559	3.0	82	0.00
40 TM	Benzene	1.370	1.333	2.7	84	0.00
41 T	Methacrylonitrile	0.289	0.271	6.2	79	0.00
42 TM	1,2-Dichloroethane	0.483	0.489	-1.2	86	0.00
43 T	Isopropyl Acetate	0.822	0.777	5.5	79	0.00
44 TM	Trichloroethene	0.411	0.393	4.4	81	0.00
45 C	1,2-Dichloropropane	0.347	0.344	0.9#	85	0.00
46 T	Dibromomethane	0.240	0.241	-0.4	86	0.00
47 T	Bromodichloromethane	0.459	0.469	-2.2	85	0.00
48 T	Methyl methacrylate	0.407	0.394	3.2	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027042.D
 Acq On : 15 Feb 2022 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 16 04:09:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	73	-0.01
50 S	Toluene-d8	1.093	1.124	-2.8	93	0.00
51 T	4-Methyl-2-Pentanone	0.498	0.497	0.2	86	0.00
52 CM	Toluene	0.867	0.863	0.5#	87	0.00
53 T	t-1,3-Dichloropropene	0.446	0.476	-6.7	85	0.00
54 T	cis-1,3-Dichloropropene	0.507	0.535	-5.5	85	0.00
55 T	1,1,2-Trichloroethane	0.328	0.348	-6.1	90	0.00
56 T	Ethyl methacrylate	0.520	0.544	-4.6	87	0.00
57 T	1,3-Dichloropropane	0.564	0.585	-3.7	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.267	0.251	6.0	74	0.00
59 T	2-Hexanone	0.364	0.362	0.5	85	0.00
60 T	Dibromochloromethane	0.338	0.362	-7.1	88	0.00
61 T	1,2-Dibromoethane	0.347	0.360	-3.7	87	0.00
62 S	4-Bromofluorobenzene	0.385	0.432	-12.2	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.502	0.456	9.2	80	0.00
65 PM	Chlorobenzene	1.034	1.024	1.0	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.377	0.8	89	0.00
67 C	Ethyl Benzene	1.830	1.848	-1.0#	91	0.00
68 T	m/p-Xylenes	0.704	0.708	-0.6	91	0.00
69 T	o-Xylene	0.690	0.709	-2.8	93	0.00
70 T	Styrene	1.119	1.178	-5.3	94	0.00
71 P	Bromoform	0.263	0.279	-6.1	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.938	3.821	3.0	92	0.00
74 T	N-amyl acetate	1.574	1.521	3.4	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.069	1.034	3.3	97	0.00
76 T	1,2,3-Trichloropropane	1.095	1.057	3.5	94	0.00
77 T	Bromobenzene	0.936	0.927	1.0	96	0.00
78 T	n-propylbenzene	4.366	4.398	-0.7	95	0.00
79 T	2-Chlorotoluene	2.698	2.631	2.5	96	0.00
80 T	1,3,5-Trimethylbenzene	3.241	3.221	0.6	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.317	0.313	1.3	90	0.00
82 T	4-Chlorotoluene	3.026	3.017	0.3	96	0.00
83 T	tert-Butylbenzene	3.089	2.985	3.4	94	0.00
84 T	1,2,4-Trimethylbenzene	3.269	3.334	-2.0	99	0.00
85 T	sec-Butylbenzene	3.916	3.945	-0.7	96	0.00
86 T	p-Isopropyltoluene	3.318	3.360	-1.3	96	0.00
87 T	1,3-Dichlorobenzene	1.715	1.711	0.2	97	0.00
88 T	1,4-Dichlorobenzene	1.730	1.717	0.8	98	0.00
89 T	n-Butylbenzene	2.714	2.839	-4.6	99	0.00
90 T	Hexachloroethane	0.479	0.519	-8.4	101	0.00
91 T	1,2-Dichlorobenzene	1.655	1.680	-1.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.263	0.242	8.0	84	0.00
93 T	1,2,4-Trichlorobenzene	1.045	1.052	-0.7	95	0.00
94 T	Hexachlorobutadiene	0.442	0.431	2.5	93	0.00
95 T	Naphthalene	3.659	3.594	1.8	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027042.D
Acq On : 15 Feb 2022 20:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 16 04:09:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.058	1.022	3.4	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6