

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027044.D
 Acq On : 16 Feb 2022 11:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 17 07:08:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	-0.01
2 T	Dichlorodifluoromethane	0.544	0.541	0.6	89	0.00
3 P	Chloromethane	0.506	0.490	3.2	89	0.00
4 C	Vinyl Chloride	0.529	0.503	4.9#	85	0.00
5 T	Bromomethane	0.225	0.201	10.7	86	0.00
6 T	Chloroethane	0.281	0.299	-6.4	87	0.00
7 T	Trichlorofluoromethane	0.779	0.760	2.4	87	0.00
8 T	Diethyl Ether	0.289	0.267	7.6	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.511	0.514	-0.6	90	0.00
10 T	Methyl Iodide	0.656	0.635	3.2	83	0.00
11 T	Tert butyl alcohol	0.108	0.076	29.6#	65	-0.02
12 CM	1,1-Dichloroethene	0.512	0.475	7.2#	83	0.00
13 T	Acrolein	0.100	0.097	3.0	87	0.00
14 T	Allyl chloride	0.906	0.873	3.6	86	0.00
15 T	Acrylonitrile	0.311	0.281	9.6	81	0.00
16 T	Acetone	0.264	0.332	-25.8#	119	-0.01
17 T	Carbon Disulfide	1.449	1.239	14.5	78	0.00
18 T	Methyl Acetate	0.758	0.648	14.5	80	0.00
19 T	Methyl tert-butyl Ether	1.811	1.652	8.8	83	0.00
20 T	Methylene Chloride	0.571	0.535	6.3	88	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.519	3.5	86	0.00
22 T	Diisopropyl ether	1.825	1.738	4.8	86	0.00
23 T	Vinyl Acetate	1.534	1.448	5.6	83	0.00
24 P	1,1-Dichloroethane	0.987	0.955	3.2	87	0.00
25 T	2-Butanone	0.425	0.429	-0.9	91	-0.01
26 T	2,2-Dichloropropane	0.667	0.717	-7.5	94	-0.01
27 T	cis-1,2-Dichloroethene	0.618	0.607	1.8	87	0.00
28 T	Bromochloromethane	0.409	0.400	2.2	88	0.00
29 T	Tetrahydrofuran	0.289	0.250	13.5	79	-0.01
30 C	Chloroform	1.026	0.979	4.6#	88	0.00
31 T	Cyclohexane	0.954	0.892	6.5	84	0.00
32 T	1,1,1-Trichloroethane	0.905	0.865	4.4	85	0.00
33 S	1,2-Dichloroethane-d4	0.600	0.579	3.5	92	-0.01
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	-0.01
35 S	Dibromofluoromethane	0.298	0.300	-0.7	93	0.00
36 T	1,1-Dichloropropene	0.470	0.469	0.2	87	0.00
37 T	Ethyl Acetate	0.534	0.475	11.0	78	-0.01
38 T	Carbon Tetrachloride	0.493	0.494	-0.2	87	0.00
39 T	Methylcyclohexane	0.576	0.590	-2.4	88	0.00
40 TM	Benzene	1.370	1.351	1.4	86	0.00
41 T	Methacrylonitrile	0.289	0.270	6.6	80	0.00
42 TM	1,2-Dichloroethane	0.483	0.490	-1.4	87	0.00
43 T	Isopropyl Acetate	0.822	0.769	6.4	79	0.00
44 TM	Trichloroethene	0.411	0.401	2.4	85	0.00
45 C	1,2-Dichloropropane	0.347	0.345	0.6#	87	0.00
46 T	Dibromomethane	0.240	0.240	0.0	87	0.00
47 T	Bromodichloromethane	0.459	0.487	-6.1	90	0.00
48 T	Methyl methacrylate	0.407	0.386	5.2	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
 Data File : VX027044.D
 Acq On : 16 Feb 2022 11:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 17 07:08:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	68	-0.01
50 S	Toluene-d8	1.093	1.118	-2.3	94	0.00
51 T	4-Methyl-2-Pentanone	0.498	0.482	3.2	85	0.00
52 CM	Toluene	0.867	0.870	-0.3#	89	0.00
53 T	t-1,3-Dichloropropene	0.446	0.492	-10.3	90	0.00
54 T	cis-1,3-Dichloropropene	0.507	0.554	-9.3	89	0.00
55 T	1,1,2-Trichloroethane	0.328	0.347	-5.8	91	0.00
56 T	Ethyl methacrylate	0.520	0.533	-2.5	87	0.00
57 T	1,3-Dichloropropane	0.564	0.583	-3.4	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.267	0.252	5.6	76	0.00
59 T	2-Hexanone	0.364	0.371	-1.9	88	0.00
60 T	Dibromochloromethane	0.338	0.381	-12.7	93	0.00
61 T	1,2-Dibromoethane	0.347	0.360	-3.7	89	0.00
62 S	4-Bromofluorobenzene	0.385	0.418	-8.6	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.502	0.468	6.8	84	0.00
65 PM	Chlorobenzene	1.034	1.033	0.1	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.384	-1.1	93	0.00
67 C	Ethyl Benzene	1.830	1.840	-0.5#	93	0.00
68 T	m/p-Xylenes	0.704	0.708	-0.6	94	0.00
69 T	o-Xylene	0.690	0.690	0.0	94	0.00
70 T	Styrene	1.119	1.163	-3.9	96	0.00
71 P	Bromoform	0.263	0.285	-8.4	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.938	3.879	1.5	95	0.00
74 T	N-amyl acetate	1.574	1.438	8.6	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.069	1.005	6.0	96	0.00
76 T	1,2,3-Trichloropropane	1.095	1.041	4.9	93	0.00
77 T	Bromobenzene	0.936	0.910	2.8	95	0.00
78 T	n-propylbenzene	4.366	4.475	-2.5	98	0.00
79 T	2-Chlorotoluene	2.698	2.650	1.8	98	0.00
80 T	1,3,5-Trimethylbenzene	3.241	3.240	0.0	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.317	0.320	-0.9	94	0.00
82 T	4-Chlorotoluene	3.026	3.015	0.4	97	0.00
83 T	tert-Butylbenzene	3.089	2.986	3.3	95	0.00
84 T	1,2,4-Trimethylbenzene	3.269	3.225	1.3	97	0.00
85 T	sec-Butylbenzene	3.916	4.029	-2.9	100	0.00
86 T	p-Isopropyltoluene	3.318	3.420	-3.1	99	0.00
87 T	1,3-Dichlorobenzene	1.715	1.711	0.2	99	0.00
88 T	1,4-Dichlorobenzene	1.730	1.711	1.1	99	0.00
89 T	n-Butylbenzene	2.714	2.981	-9.8	105	0.00
90 T	Hexachloroethane	0.479	0.545	-13.8	107	0.00
91 T	1,2-Dichlorobenzene	1.655	1.629	1.6	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.263	0.237	9.9	84	0.00
93 T	1,2,4-Trichlorobenzene	1.045	1.093	-4.6	100	0.00
94 T	Hexachlorobutadiene	0.442	0.468	-5.9	103	0.00
95 T	Naphthalene	3.659	3.516	3.9	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021622\
Data File : VX027044.D
Acq On : 16 Feb 2022 11:20
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 17 07:08:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.058	1.051	0.7	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6