

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040267.D
 Acq On : 16 Feb 2024 10:56
 Operator : JC/MD
 Sample : P1507-07 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D7-20240214

Quant Time: Feb 17 01:31:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73724	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	126571	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50378	45.932	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.860%
35) Dibromofluoromethane	5.385	113	41063	46.456	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.920%
50) Toluene-d8	8.647	98	155081	46.903	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.800%
62) 4-Bromofluorobenzene	11.079	95	59974	44.475	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.960%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	1061	1.357	ug/l	98
44) Trichloroethene	7.122	130	62530	65.468	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040267.D
Acq On : 16 Feb 2024 10:56
Operator : JC/MD
Sample : P1507-07 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D7-20240214

Quant Time: Feb 17 01:31:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

