

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040273.D
 Acq On : 16 Feb 2024 13:13
 Operator : JC/MD
 Sample : P1507-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20240212

Quant Time: Feb 17 01:34:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71121	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	121180	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48395	45.739	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.480%
35) Dibromofluoromethane	5.385	113	40109	47.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.647	98	148553	46.927	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.860%
62) 4-Bromofluorobenzene	11.079	95	59316	45.944	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.880%
Target Compounds						
					Qvalue	
30) Chloroform	5.093	83	729	0.453	ug/l	# 82
44) Trichloroethene	7.129	130	3872	4.234	ug/l	98
64) Tetrachloroethene	9.275	164	646	0.770	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040273.D
Acq On : 16 Feb 2024 13:13
Operator : JC/MD
Sample : P1507-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20240212

Quant Time: Feb 17 01:34:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

