

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040275.D
 Acq On : 16 Feb 2024 13:59
 Operator : JC/MD
 Sample : P1507-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20240214

Quant Time: Feb 17 03:36:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71630	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	121091	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49267	46.232	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.460%
35) Dibromofluoromethane	5.385	113	40028	47.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.660%
50) Toluene-d8	8.647	98	148347	46.897	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.800%
62) 4-Bromofluorobenzene	11.079	95	57296	44.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.820%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	4128	5.435	ug/l	90
12) 1,1-Dichloroethene	2.325	96	569	0.778	ug/l	79
27) cis-1,2-Dichloroethene	4.495	96	1297	1.365	ug/l	81
44) Trichloroethene	7.123	130	163918	179.387	ug/l	98
64) Tetrachloroethene	9.275	164	4590	5.513	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040275.D
Acq On : 16 Feb 2024 13:59
Operator : JC/MD
Sample : P1507-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20240214

Quant Time: Feb 17 03:36:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

