

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040278.D
 Acq On : 16 Feb 2024 15:08
 Operator : JC/MD
 Sample : P1507-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D3-20240214

Quant Time: Feb 17 01:36:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119131	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48944	45.989	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.980%
35) Dibromofluoromethane	5.385	113	39923	47.138	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	8.647	98	149084	47.057	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.120%
62) 4-Bromofluorobenzene	11.079	95	57400	44.424	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.840%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	381	0.471	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	2593	3.418	ug/l	96
12) 1,1-Dichloroethene	2.318	96	252	0.345	ug/l #	79
27) cis-1,2-Dichloroethene	4.489	96	696	0.733	ug/l	70
44) Trichloroethene	7.123	130	65381	71.440	ug/l	92
64) Tetrachloroethene	9.275	164	1951	2.337	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040278.D
Acq On : 16 Feb 2024 15:08
Operator : JC/MD
Sample : P1507-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D3-20240214

Quant Time: Feb 17 01:36:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

