

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040279.D
 Acq On : 16 Feb 2024 15:31
 Operator : JC/MD
 Sample : P1507-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D2-20240214

Quant Time: Feb 17 01:37:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70551	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	119785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	48308	46.026	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.060%
35) Dibromofluoromethane	5.379	113	39370	47.064	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.120%
50) Toluene-d8	8.647	98	145198	46.402	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.800%
62) 4-Bromofluorobenzene	11.079	95	57414	44.989	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.980%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	1197	1.502	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.337	101	2986	3.991	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	268	0.286	ug/l	63
44) Trichloroethene	7.135	130	3042	3.365	ug/l	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040279.D
Acq On : 16 Feb 2024 15:31
Operator : JC/MD
Sample : P1507-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D2-20240214

Quant Time: Feb 17 01:37:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

