

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040282.D
 Acq On : 16 Feb 2024 16:39
 Operator : JC/MD
 Sample : P1507-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D1-20240215

Quant Time: Feb 17 01:38:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70238	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	117287	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	47447	45.407	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.820%
35) Dibromofluoromethane	5.385	113	38436	46.926	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.860%
50) Toluene-d8	8.647	98	144234	47.075	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.160%
62) 4-Bromofluorobenzene	11.079	95	56860	45.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.000%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	508	0.682	ug/l	# 92
12) 1,1-Dichloroethene	2.319	96	440	0.614	ug/l	# 71
19) Methyl tert-butyl Ether	3.117	73	1592	0.630	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	11995	12.874	ug/l	80
44) Trichloroethene	7.123	130	96213	108.708	ug/l	99
64) Tetrachloroethene	9.275	164	11023	13.720	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040282.D
Acq On : 16 Feb 2024 16:39
Operator : JC/MD
Sample : P1507-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D1-20240215

Quant Time: Feb 17 01:38:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

