

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040285.D
 Acq On : 16 Feb 2024 17:48
 Operator : JC/MD
 Sample : P1507-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D1-20240215

Quant Time: Feb 17 01:39:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67518	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	116624	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	46987	46.778	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.560%
35) Dibromofluoromethane	5.379	113	38038	46.705	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.400%
50) Toluene-d8	8.646	98	141649	46.495	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.980%
62) 4-Bromofluorobenzene	11.079	95	55040	44.297	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.600%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.128	130	13441	15.273	ug/l	94
64) Tetrachloroethene	9.274	164	628	0.789	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040285.D
Acq On : 16 Feb 2024 17:48
Operator : JC/MD
Sample : P1507-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D1-20240215

Quant Time: Feb 17 01:39:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

