

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040262.D
 Acq On : 16 Feb 2024 08:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 17 01:28:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.565	0.616	-9.0	101	0.00
3 P	Chloromethane	0.582	0.550	5.5	96	0.00
4 C	Vinyl Chloride	0.615	0.610	0.8#	95	0.00
5 T	Bromomethane	0.347	0.438	-26.2#	109	0.00
6 T	Chloroethane	0.388	0.375	3.4	96	0.00
7 T	Trichlorofluoromethane	0.988	0.991	-0.3	98	0.00
8 T	Diethyl Ether	0.371	0.348	6.2	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.530	0.563	-6.2	105	0.00
10 T	Methyl Iodide	0.589	0.681	-15.6	108	0.00
11 T	Tert butyl alcohol	0.145	0.105	27.6#	72	-0.01
12 CM	1,1-Dichloroethene	0.510	0.525	-2.9#	102	0.00
13 T	Acrolein	0.134	0.113	15.7	89	0.00
14 T	Allyl chloride	0.886	0.863	2.6	97	0.00
15 T	Acrylonitrile	0.324	0.314	3.1	94	0.00
16 T	Acetone	0.306	0.318	-3.9	105	0.00
17 T	Carbon Disulfide	1.348	1.278	5.2	99	0.00
18 T	Methyl Acetate	0.819	0.684	16.5	82	0.00
19 T	Methyl tert-butyl Ether	1.798	1.814	-0.9	97	0.00
20 T	Methylene Chloride	0.624	0.597	4.3	101	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.566	2.6	101	0.00
22 T	Diisopropyl ether	1.797	1.837	-2.2	98	-0.01
23 T	Vinyl Acetate	1.602	1.644	-2.6	98	0.00
24 P	1,1-Dichloroethane	1.047	1.044	0.3	99	0.00
25 T	2-Butanone	0.478	0.446	6.7	91	-0.01
26 T	2,2-Dichloropropane	0.806	0.857	-6.3	103	0.00
27 T	cis-1,2-Dichloroethene	0.663	0.665	-0.3	101	0.00
28 T	Bromochloromethane	0.444	0.470	-5.9	100	0.00
29 T	Tetrahydrofuran	0.311	0.285	8.4	90	-0.02
30 C	Chloroform	1.132	1.140	-0.7#	99	-0.01
31 T	Cyclohexane	0.887	0.860	3.0	98	-0.01
32 T	1,1,1-Trichloroethane	0.952	0.984	-3.4	100	-0.01
33 S	1,2-Dichloroethane-d4	0.744	0.697	6.3	95	-0.01
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.349	0.358	-2.6	100	0.00
36 T	1,1-Dichloropropene	0.489	0.512	-4.7	100	-0.01
37 T	Ethyl Acetate	0.544	0.525	3.5	94	-0.02
38 T	Carbon Tetrachloride	0.465	0.494	-6.2	100	0.00
39 T	Methylcyclohexane	0.568	0.595	-4.8	99	0.00
40 TM	Benzene	1.368	1.441	-5.3	99	0.00
41 T	Methacrylonitrile	0.289	0.294	-1.7	96	-0.02
42 TM	1,2-Dichloroethane	0.535	0.557	-4.1	95	0.00
43 T	Isopropyl Acetate	0.850	0.870	-2.4	95	-0.01
44 TM	Trichloroethene	0.377	0.400	-6.1	101	0.00
45 C	1,2-Dichloropropane	0.356	0.394	-10.7#	103	0.00
46 T	Dibromomethane	0.262	0.287	-9.5	99	0.00
47 T	Bromodichloromethane	0.505	0.558	-10.5	102	0.00
48 T	Methyl methacrylate	0.435	0.442	-1.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040262.D
 Acq On : 16 Feb 2024 08:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 17 01:28:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	84	0.00
50 S	Toluene-d8	1.306	1.320	-1.1	97	0.00
51 T	4-Methyl-2-Pentanone	0.580	0.591	-1.9	93	0.00
52 CM	Toluene	0.896	0.953	-6.4#	98	0.00
53 T	t-1,3-Dichloropropene	0.528	0.598	-13.3	100	0.00
54 T	cis-1,3-Dichloropropene	0.566	0.624	-10.2	99	0.00
55 T	1,1,2-Trichloroethane	0.382	0.412	-7.9	103	0.00
56 T	Ethyl methacrylate	0.572	0.615	-7.5	95	0.00
57 T	1,3-Dichloropropane	0.624	0.676	-8.3	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.271	0.278	-2.6	93	0.00
59 T	2-Hexanone	0.460	0.474	-3.0	93	0.00
60 T	Dibromochloromethane	0.389	0.434	-11.6	99	0.00
61 T	1,2-Dibromoethane	0.406	0.445	-9.6	101	0.00
62 S	4-Bromofluorobenzene	0.533	0.546	-2.4	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.350	0.378	-8.0	107	0.00
65 PM	Chlorobenzene	1.044	1.111	-6.4	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.410	-13.6	102	0.00
67 C	Ethyl Benzene	1.859	2.040	-9.7#	100	0.00
68 T	m/p-Xylenes	0.702	0.775	-10.4	102	0.00
69 T	o-Xylene	0.682	0.757	-11.0	100	0.00
70 T	Styrene	1.134	1.280	-12.9	101	0.00
71 P	Bromoform	0.296	0.341	-15.2	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.522	3.728	-5.8	103	0.00
74 T	N-amyl acetate	1.510	1.585	-5.0	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.265	1.279	-1.1	99	0.00
76 T	1,2,3-Trichloropropane	1.013	1.063	-4.9	99	0.00
77 T	Bromobenzene	0.832	0.894	-7.5	105	0.00
78 T	n-propylbenzene	4.264	4.524	-6.1	102	0.00
79 T	2-Chlorotoluene	2.575	2.620	-1.7	100	0.00
80 T	1,3,5-Trimethylbenzene	2.965	3.172	-7.0	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.328	0.350	-6.7	101	0.00
82 T	4-Chlorotoluene	3.050	3.133	-2.7	101	0.00
83 T	tert-Butylbenzene	2.907	3.051	-5.0	102	0.00
84 T	1,2,4-Trimethylbenzene	3.000	3.218	-7.3	103	0.00
85 T	sec-Butylbenzene	3.736	4.050	-8.4	104	0.00
86 T	p-Isopropyltoluene	3.062	3.348	-9.3	103	0.00
87 T	1,3-Dichlorobenzene	1.648	1.760	-6.8	103	0.00
88 T	1,4-Dichlorobenzene	1.745	1.800	-3.2	104	0.00
89 T	n-Butylbenzene	2.956	3.203	-8.4	104	0.00
90 T	Hexachloroethane	0.483	0.542	-12.2	103	0.00
91 T	1,2-Dichlorobenzene	1.595	1.754	-10.0	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.292	0.299	-2.4	95	0.00
93 T	1,2,4-Trichlorobenzene	1.092	1.160	-6.2	105	0.00
94 T	Hexachlorobutadiene	0.436	0.461	-5.7	109	0.00
95 T	Naphthalene	3.450	3.621	-5.0	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040262.D
Acq On : 16 Feb 2024 08:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 17 01:28:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.041	1.141	-9.6	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6