

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040270.D
 Acq On : 16 Feb 2024 12:04
 Operator : JC/MD
 Sample : P1507-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20240212

Quant Time: Feb 17 01:33:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74330	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	125948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51880	46.916	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.840%
35) Dibromofluoromethane	5.391	113	41431	47.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.647	98	155425	47.240	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.480%
62) 4-Bromofluorobenzene	11.079	95	60346	44.972	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.940%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040270.D
Acq On : 16 Feb 2024 12:04
Operator : JC/MD
Sample : P1507-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20240212

Quant Time: Feb 17 01:33:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX040270.D
Acq: 16 Feb 2024 12:04




Instrument :
MSVOA_X
ClientSampleId :
TB01-20240212


Tgt Ion:168 Resp: 74330
Ion Ratio Lower Upper
168 100
99 56.2 56.6 85.0#



#33
1,2-Dichloroethane-d4
Concen: 46.916 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.000 min
Lab File: VX040270.D
Acq: 16 Feb 2024 12:04

Tgt Ion: 65 Resp: 51880
Ion Ratio Lower Upper
65 100
67 51.8 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX040270.D
Acq: 16 Feb 2024 12:04




Instrument :
MSVOA_X
ClientSampleId :
TB01-20240212


Tgt Ion:114 Resp: 125948
Ion Ratio Lower Upper
114 100
63 19.8 0.0 45.0
88 15.0 0.0 36.6



#35
Dibromofluoromethane
Concen: 47.105 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.006 min
Lab File: VX040270.D
Acq: 16 Feb 2024 12:04

Tgt Ion:113 Resp: 41431
Ion Ratio Lower Upper
113 100
111 102.9 82.9 124.3
192 18.7 13.3 19.9









#50
Toluene-d8
Concen: 47.240 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040270.D
Acq: 16 Feb 2024 12:04




Instrument :
MSVOA_X
ClientSampleId :
TB01-20240212


Tgt Ion: 98 Resp: 155425
Ion Ratio Lower Upper
98 100
100 64.0 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 44.972 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX040270.D
Acq: 16 Feb 2024 12:04

Tgt Ion: 95 Resp: 60346
Ion Ratio Lower Upper
95 100
174 77.8 0.0 135.6
176 75.0 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX040270.D
Acq: 16 Feb 2024 12:04

Instrument :
MSVOA_X
ClientSampleId :
TB01-20240212

Tgt Ion:117 Resp: 122477
Ion Ratio Lower Upper
117 100
82 56.9 46.2 69.4
119 32.2 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX040270.D
Acq: 16 Feb 2024 12:04

Tgt Ion:152 Resp: 62976
Ion Ratio Lower Upper
152 100
115 60.7 43.3 129.9
150 157.8 0.0 345.0

