

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040288.D
 Acq On : 16 Feb 2024 18:57
 Operator : JC/MD
 Sample : P1507-30MSD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20240212MSD

Quant Time: Feb 17 01:41:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72875	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	125757	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	53087	48.966	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.940%
35) Dibromofluoromethane	5.391	113	43170	49.156	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.320%
50) Toluene-d8	8.647	98	156682	47.694	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.380%
62) 4-Bromofluorobenzene	11.079	95	65479	48.872	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.740%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	40338	49.002	ug/l	97
3) Chloromethane	1.294	50	43881	51.693	ug/l	94
4) Vinyl Chloride	1.374	62	46126	51.419	ug/l	99
5) Bromomethane	1.599	94	28744	56.782	ug/l	97
6) Chloroethane	1.678	64	29698	52.493	ug/l	98
7) Trichlorofluoromethane	1.886	101	73010	50.680	ug/l	99
8) Diethyl Ether	2.130	74	28994	53.577	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.325	101	714289	924.346	ug/l	98
10) Methyl Iodide	2.453	142	51433	59.951	ug/l	93
11) Tert butyl alcohol	2.965	59	68451	324.138	ug/l	# 93
12) 1,1-Dichloroethene	2.318	96	47017	63.193	ug/l	86
13) Acrolein	2.233	56	50794	260.589	ug/l	99
14) Allyl chloride	2.660	41	69226	53.615	ug/l	97
15) Acrylonitrile	3.062	53	139525	295.286	ug/l	98
16) Acetone	2.379	43	116459	261.544	ug/l	93
17) Carbon Disulfide	2.507	76	96536	49.146	ug/l	100
18) Methyl Acetate	2.703	43	56822	47.603	ug/l	95
19) Methyl tert-butyl Ether	3.111	73	150977	57.598	ug/l	97
20) Methylene Chloride	2.788	84	48586	58.474	ug/l	96
21) trans-1,2-Dichloroethene	3.093	96	44290	52.268	ug/l	93
22) Diisopropyl ether	3.757	45	152549	58.256	ug/l	96
23) Vinyl Acetate	3.721	43	662192	283.671	ug/l	98
24) 1,1-Dichloroethane	3.611	63	86558	56.747	ug/l	99
25) 2-Butanone	4.556	43	196935	282.942	ug/l	98
26) 2,2-Dichloropropane	4.471	77	56164	47.832	ug/l	94
27) cis-1,2-Dichloroethene	4.489	96	58671	60.694	ug/l	92
28) Bromochloromethane	4.897	49	35794	55.271	ug/l	92
29) Tetrahydrofuran	5.001	42	133840	295.155	ug/l	96
30) Chloroform	5.092	83	94585	57.353	ug/l	99
31) Cyclohexane	5.464	56	63186	48.862	ug/l	98
32) 1,1,1-Trichloroethane	5.385	97	78053	56.259	ug/l	99
36) 1,1-Dichloropropene	5.696	75	65273	53.057	ug/l	99
37) Ethyl Acetate	4.714	43	72139	52.699	ug/l	99
38) Carbon Tetrachloride	5.678	117	62729	53.628	ug/l	98
39) Methylcyclohexane	7.379	83	63208	44.211	ug/l	94
40) Benzene	6.037	78	189022	54.944	ug/l	99

02/17/2024
 Supervised By :Semsettin
 Cesilyurt

02/20/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
 Data File : VX040288.D
 Acq On : 16 Feb 2024 18:57
 Operator : JC/MD
 Sample : P1507-30MSD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20240212MSD

Quant Time: Feb 17 01:41:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	42217	58.176	ug/l	93
42) 1,2-Dichloroethane	6.086	62	75175	55.904	ug/l	98
43) Isopropyl Acetate	6.342	43	117494	54.984	ug/l	98
44) Trichloroethene	7.123	130	226882	239.081	ug/l	96
45) 1,2-Dichloropropane	7.433	63	52400	58.518	ug/l	97
46) Dibromomethane	7.580	93	37934	57.520	ug/l	98
47) Bromodichloromethane	7.824	83	73539	57.872	ug/l	100
48) Methyl methacrylate	7.689	41	62206	56.892	ug/l	94
49) 1,4-Dioxane	7.659	88	29090	1222.951	ug/l #	92
51) 4-Methyl-2-Pentanone	8.573	43	431827	295.986	ug/l	100
52) Toluene	8.720	92	124234	55.131	ug/l	94
53) t-1,3-Dichloropropene	8.976	75	77196	51.025	ug/l	96
54) cis-1,3-Dichloropropene	8.366	75	78951	55.438	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	55070	57.298	ug/l	94
56) Ethyl methacrylate	9.116	69	83584	58.091	ug/l	96
57) 1,3-Dichloropropane	9.305	76	88542	56.441	ug/l	99
59) 2-Hexanone	9.427	43	336447	290.862	ug/l	97
60) Dibromochloromethane	9.518	129	56616	57.818	ug/l	99
61) 1,2-Dibromoethane	9.610	107	56946	55.741	ug/l	99
64) Tetrachloroethene	9.275	164	293415	341.784	ug/l	99
65) Chlorobenzene	10.079	112	135996	53.191	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	49648	56.067	ug/l	98
67) Ethyl Benzene	10.195	91	251131	55.150	ug/l	98
68) m/p-Xylenes	10.299	106	188012	109.324	ug/l	98
69) o-Xylene	10.640	106	95348	57.053	ug/l	99
70) Styrene	10.652	104	160611	57.805	ug/l	99
71) Bromoform	10.799	173	43163	51.837	ug/l #	100
73) Isopropylbenzene	10.963	105	253251	54.069	ug/l	99
74) N-amyl acetate	10.841	43	110661	55.118	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.213	83	94922	56.408	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	77427m	57.495	ug/l	
77) Bromobenzene	11.195	156	63072	57.013	ug/l	94
78) n-propylbenzene	11.305	91	309923	54.654	ug/l	99
79) 2-Chlorotoluene	11.366	91	183568	53.608	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	215906	54.753	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	23500	49.074	ug/l	95
82) 4-Chlorotoluene	11.451	91	217960	53.722	ug/l	99
83) tert-Butylbenzene	11.713	119	209914	54.298	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	219996	55.145	ug/l	100
85) sec-Butylbenzene	11.890	105	271255	54.596	ug/l	100
86) p-Isopropyltoluene	12.006	119	222957	54.745	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	120055	54.771	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	122872	52.938	ug/l	99
89) n-Butylbenzene	12.335	91	210512	53.547	ug/l	97
90) Hexachloroethane	12.536	117	37218	57.935	ug/l	90
91) 1,2-Dichlorobenzene	12.335	146	117152	55.210	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	23325	60.056	ug/l	92
93) 1,2,4-Trichlorobenzene	13.585	180	77240	53.169	ug/l	96
94) Hexachlorobutadiene	13.725	225	28803	49.678	ug/l	98
95) Naphthalene	13.774	128	258428	56.324	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	75398	54.450	ug/l	97

02/17/2024
 Supervised By :Semsettin
 Yesilyurt

02/20/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040288.D
Acq On : 16 Feb 2024 18:57
Operator : JC/MD
Sample : P1507-30MSD
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Quant Time: Feb 17 01:41:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD

Manual Integrations
APPROVED

Reviewed By :John
Carlone

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#)=qualifier out of range (m)=manual integration (+)=signals summed						

02/17/2024
Supervised By :Semsettin
Yesilyurt

02/20/2024


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040288.D
Acq On : 16 Feb 2024 18:57
Operator : JC/MD
Sample : P1507-30MSD
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD

Quant Time: Feb 17 01:41:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :John
Carlone





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:168 Resp: 72875
Ion Ratio Lower Upper
168 100
99 55.0 56.6 85.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#2
Dichlorodifluoromethane
Concen: 49.002 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 85 Resp: 40338
Ion Ratio Lower Upper
85 100
87 35.3 16.7 50.1





#3
Chloromethane
Concen: 51.693 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 50 Resp: 4388
Ion Ratio Lower Upper
50 100
52 35.5 25.5 38.3


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#4
Vinyl Chloride
Concen: 51.419 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 62 Resp: 46126
Ion Ratio Lower Upper
62 100
64 31.1 25.4 38.0





#5
Bromomethane
Concen: 56.782 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 94 Resp: 28744
Ion Ratio Lower Upper
94 100
96 96.0 74.6 111.8


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#6
Chloroethane
Concen: 52.493 ug/l
RT: 1.678 min Scan# 97
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 64 Resp: 29698
Ion Ratio Lower Upper
64 100
66 32.7 25.1 37.7





#7
Trichlorofluoromethane
Concen: 50.680 ug/l
RT: 1.886 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:101 Resp: 73010
Ion Ratio Lower Upper
101 100
103 65.4 51.4 77.2


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#8
Diethyl Ether
Concen: 53.577 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion: 74 Resp: 28994
Ion Ratio Lower Upper
74 100
45 95.5 53.3 159.8





#9
1,1,2-Trichlorotrifluoroethane
Concen: 924.346 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:101 Resp: 714289
Ion Ratio Lower Upper
101 100
85 46.5 37.0 55.6
151 75.5 58.6 87.8


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#10
Methyl Iodide
Concen: 59.951 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion:142 Resp: 51433
Ion Ratio Lower Upper
142 100
127 45.6 41.6 62.4
141 14.3 11.6 17.4





#11
Tert butyl alcohol
Concen: 324.138 ug/l
RT: 2.965 min Scan# 308
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 59 Resp: 6845
Ion Ratio Lower Upper
59 100
57 9.8 6.0 9.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#12
1,1-Dichloroethene
Concen: 63.193 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 96 Resp: 47017
Ion Ratio Lower Upper
96 100
61 155.3 144.6 216.8
98 61.4 50.4 75.6





#13
Acrolein
Concen: 260.589 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 56 Resp: 50794
Ion Ratio Lower Upper
56 100
55 72.6 57.1 85.7


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#14
Allyl chloride
Concen: 53.615 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion: 41 Resp: 69226
Ion Ratio Lower Upper
41 100
39 70.7 55.8 83.6
76 35.0 24.3 36.5





#15
Acrylonitrile
Concen: 295.286 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 53 Resp: 139529
Ion Ratio Lower Upper
53 100
52 83.9 66.0 99.0
51 36.4 29.8 44.6



Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#16
Acetone
Concen: 261.544 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Tgt Ion: 43 Resp: 116459
Ion Ratio Lower Upper
43 100
58 31.1 21.9 32.9





#17

Carbon Disulfide

Concen: 49.146 ug/l

RT: 2.507 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX040288.D

Acq: 16 Feb 2024 18:57

Instrument :

MSVOA_X

ClientSampleId :

RE134D3-20240212MSD

Tgt Ion: 76 Resp: 96530

Ion Ratio Lower Upper

76 100



78 8.8 7.0 10.6

Manual Integrations

APPROVED

Reviewed By :John

Carlone





02/17/2024

Supervised By :Semsettin

Yesilyurt

02/20/2024



#18

Methyl Acetate

Concen: 47.603 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX040288.D




Acq: 16 Feb 2024 18:57


Tgt Ion: 43 Resp: 56822

Ion Ratio Lower Upper

43 100


74 24.5 17.8 26.8





#19
Methyl tert-butyl Ether
Concen: 57.598 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 73 Resp: 15097
Ion Ratio Lower Upper
73 100
57 20.4 17.3 25.9


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#20
Methylene Chloride
Concen: 58.474 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 84 Resp: 48586
Ion Ratio Lower Upper
84 100
49 125.2 103.6 155.4
51 37.8 33.8 50.6
86 63.8 49.4 74.0





#21
trans-1,2-Dichloroethene
Concen: 52.268 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 96 Resp: 44290
Ion Ratio Lower Upper
96 100
61 139.2 119.2 178.8
98 65.0 48.7 73.1


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#22
Diisopropyl ether
Concen: 58.256 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 45 Resp: 152549
Ion Ratio Lower Upper
45 100
43 91.2 69.7 104.5
87 27.4 20.8 31.2
59 12.6 9.0 13.4





#23
Vinyl Acetate
Concen: 283.671 ug/l
RT: 3.721 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 43 Resp: 662192
Ion Ratio Lower Upper
43 100
86 10.6 7.8 11.8


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#24
1,1-Dichloroethane
Concen: 56.747 ug/l
RT: 3.611 min Scan# 414
Delta R.T. 0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 63 Resp: 86558
Ion Ratio Lower Upper
63 100
98 7.1 3.3 9.8
100 4.2 2.1 6.5





#25
2-Butanone
Concen: 282.942 ug/l
RT: 4.556 min Scan# 50
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 43 Resp: 19693
Ion Ratio Lower Upper
43 100
72 25.4 19.4 29.2


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#26
2,2-Dichloropropane
Concen: 47.832 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 77 Resp: 56164
Ion Ratio Lower Upper
77 100
97 24.2 10.8 32.3





#27
cis-1,2-Dichloroethene
Concen: 60.694 ug/l
RT: 4.489 min Scan# 558
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 96 Resp: 5867
Ion Ratio Lower Upper
96 100
61 140.9 0.0 312.2
98 64.9 0.0 129.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#28
Bromochloromethane
Concen: 55.271 ug/l
RT: 4.897 min Scan# 625
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 49 Resp: 35794
Ion Ratio Lower Upper
49 100
129 2.0 0.0 3.8
130 79.2 57.5 86.3





#29
Tetrahydrofuran
Concen: 295.155 ug/l
RT: 5.001 min Scan# 642
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument : MSVOA_X
ClientSampleId : RE134D3-20240212MSD



Tgt Ion: 42 Resp: 133840
Ion Ratio Lower Upper
42 100
72 45.7 34.2 51.4
71 41.9 32.2 48.2


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#30
Chloroform
Concen: 57.353 ug/l
RT: 5.092 min Scan# 657
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 83 Resp: 94585
Ion Ratio Lower Upper
83 100
85 61.9 50.3 75.5





#31
Cyclohexane
Concen: 48.862 ug/l
RT: 5.464 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 56 Resp: 63180
Ion Ratio Lower Upper
56 100
69 28.3 23.4 35.2
84 84.2 65.9 98.9


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#32
1,1,1-Trichloroethane
Concen: 56.259 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Tgt Ion: 97 Resp: 78053
Ion Ratio Lower Upper
97 100
99 64.3 51.1 76.7
61 47.2 36.4 54.6







#33
1,2-Dichloroethane-d4
Concen: 48.966 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD

Tgt Ion: 65 Resp: 5308
Ion Ratio Lower Upper
65 100
67 52.5 0.0 100.2



Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion:114 Resp: 125757
Ion Ratio Lower Upper
114 100
63 20.3 0.0 45.0
88 15.3 0.0 36.6





#35
Dibromofluoromethane
Concen: 49.156 ug/l
RT: 5.391 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:113 Resp: 43170
Ion Ratio Lower Upper
113 100
111 98.7 82.9 124.3
192 18.1 13.3 19.9


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#36
1,1-Dichloropropene
Concen: 53.057 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 75 Resp: 65273
Ion Ratio Lower Upper
75 100
110 35.0 17.0 50.9
77 30.8 24.7 37.1





#37
Ethyl Acetate
Concen: 52.699 ug/l
RT: 4.714 min Scan# 595
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 43 Resp: 72139
Ion Ratio Lower Upper
43 100
61 14.0 10.6 16.0
70 10.8 8.3 12.5


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#38
Carbon Tetrachloride
Concen: 53.628 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion:117 Resp: 62729
Ion Ratio Lower Upper
117 100
119 95.8 78.2 117.2
121 28.7 24.4 36.6





#39
Methylcyclohexane
Concen: 44.211 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 83 Resp: 63208
Ion Ratio Lower Upper
83 100
55 73.4 64.6 97.0
98 43.9 36.7 55.1


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#40
Benzene
Concen: 54.944 ug/l
RT: 6.037 min Scan# 812
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 78 Resp: 189022
Ion Ratio Lower Upper
78 100
77 23.5 19.3 28.9





#41
Methacrylonitrile
Concen: 58.176 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.012 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 41 Resp: 4221
Ion Ratio Lower Upper
41 100
39 57.9 45.2 67.8
67 75.0 52.6 78.8
52 32.0 22.7 34.1


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#42
1,2-Dichloroethane
Concen: 55.904 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 62 Resp: 75175
Ion Ratio Lower Upper
62 100
98 8.6 0.0 16.0





#43
Isopropyl Acetate
Concen: 54.984 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 43 Resp: 117494
Ion Ratio Lower Upper
43 100
61 19.3 15.2 22.8
87 13.7 9.8 14.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#44
Trichloroethene
Concen: 239.081 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Tgt Ion:130 Resp: 226882
Ion Ratio Lower Upper
130 100
95 106.3 0.0 204.4





#45

1,2-Dichloropropane

Concen: 58.518 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VX040288.D

Acq: 16 Feb 2024 18:57

Instrument :

MSVOA_X

ClientSampleId :

RE134D3-20240212MSD

Tgt Ion: 63 Resp: 52400

Ion Ratio Lower Upper



63 100

65 32.3 24.3 36.5

Manual Integrations



APPROVED

Reviewed By :John
Carlone



02/17/2024
Supervised By :Semsettin
Yesilyurt

02/20/2024



#46

Dibromomethane

Concen: 57.520 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX040288.D

Acq: 16 Feb 2024 18:57




Tgt Ion: 93 Resp: 37934


Ion Ratio Lower Upper

93 100

95 83.4 66.6 100.0


174 91.0 69.4 104.0





#47
Bromodichloromethane
Concen: 57.872 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 83 Resp: 73539
Ion Ratio Lower Upper
83 100
85 62.1 49.8 74.6
127 8.7 7.1 10.7


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#48
Methyl methacrylate
Concen: 56.892 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 41 Resp: 62206
Ion Ratio Lower Upper
41 100
69 82.0 59.2 88.8
39 61.7 47.7 71.5





#49
1,4-Dioxane
Concen: 1222.951 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 88 Resp: 29090
Ion Ratio Lower Upper
88 100
43 33.3 0.0 0.0
58 68.0 60.1 90.1


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#50
Toluene-d8
Concen: 47.694 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Tgt Ion: 98 Resp: 156682
Ion Ratio Lower Upper
98 100
100 63.9 53.1 79.7







#51
4-Methyl-2-Pentanone
Concen: 295.986 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.001 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD

Tgt Ion: 43 Resp: 43182
Ion Ratio Lower Upper
43 100
58 38.0 30.4 45.6



Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#52
Toluene
Concen: 55.131 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Tgt Ion: 92 Resp: 124234
Ion Ratio Lower Upper
92 100
91 175.3 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 51.025 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX040288.D

Acq: 16 Feb 2024 18:57

Instrument :

MSVOA_X

ClientSampleId :

RE134D3-20240212MSD

Tgt Ion: 75 Resp: 77190

Ion Ratio Lower Upper

75 100



77 30.1 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :John

Carlone





02/17/2024

Supervised By :Semsettin

Yesilyurt

02/20/2024



#54

cis-1,3-Dichloropropene

Concen: 55.438 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX040288.D

Acq: 16 Feb 2024 18:57




Tgt Ion: 75 Resp: 78951


Ion Ratio Lower Upper

75 100

77 31.4 25.3 37.9


39 53.0 46.7 70.1





#55
1,1,2-Trichloroethane
Concen: 57.298 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD





Tgt Ion: 97 Resp: 55070

Ion	Ratio	Lower	Upper
97	100		
83	88.5	65.8	98.8
85	56.6	41.6	62.4
99	64.0	47.8	71.8


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt

02/20/2024






#56
Ethyl methacrylate
Concen: 58.091 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Tgt Ion: 69 Resp: 83584

Ion	Ratio	Lower	Upper
69	100		
41	68.0	56.8	85.2
39	42.7	32.0	48.0







#57
1,3-Dichloropropane
Concen: 56.441 ug/l
RT: 9.305 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD

Tgt Ion: 76 Resp: 8854
Ion Ratio Lower Upper
76 100
78 32.1 26.1 39.1



Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#59
2-Hexanone
Concen: 290.862 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Tgt Ion: 43 Resp: 336447
Ion Ratio Lower Upper
43 100
58 54.3 26.2 78.5





#60

Dibromochloromethane

Concen: 57.818 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX040288.D

Acq: 16 Feb 2024 18:57

Instrument :

MSVOA_X

ClientSampleId :

RE134D3-20240212MSD

Tgt Ion:129 Resp: 56610

Ion Ratio Lower Upper



129 100

127 78.3 38.8 116.4

Manual Integrations



APPROVED

Reviewed By :John
Carlone



02/17/2024
Supervised By :Semsettin
Yesilyurt

02/20/2024



#61

1,2-Dibromoethane

Concen: 55.741 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. -0.000 min

Lab File: VX040288.D




Acq: 16 Feb 2024 18:57


Tgt Ion:107 Resp: 56946

Ion Ratio Lower Upper

107 100


109 95.0 75.2 112.8





#62
4-Bromofluorobenzene
Concen: 48.872 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 95 Resp: 65479
Ion Ratio Lower Upper
95 100
174 76.4 0.0 135.6
176 73.4 0.0 131.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion:117 Resp: 122506
Ion Ratio Lower Upper
117 100
82 56.5 46.2 69.4
119 30.2 25.5 38.3





#64
Tetrachloroethene
Concen: 341.784 ug/l
RT: 9.275 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:164 Resp: 29341
Ion Ratio Lower Upper
164 100
166 126.6 102.7 154.1
129 99.4 80.4 120.6
131 96.0 76.5 114.7


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#65
Chlorobenzene
Concen: 53.191 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion:112 Resp: 135996
Ion Ratio Lower Upper
112 100
114 33.2 25.7 38.5





#66
1,1,1,2-Tetrachloroethane
Concen: 56.067 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:131 Resp: 49648
Ion Ratio Lower Upper
131 100
133 95.6 47.1 141.4
119 67.0 32.2 96.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#67
Ethyl Benzene
Concen: 55.150 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 91 Resp: 251131
Ion Ratio Lower Upper
91 100
106 30.2 24.9 37.3





#68
m/p-Xylenes
Concen: 109.324 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:106 Resp: 188012
Ion Ratio Lower Upper
106 100
91 210.2 165.5 248.3


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#69
o-Xylene
Concen: 57.053 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion:106 Resp: 95348
Ion Ratio Lower Upper
106 100
91 216.3 108.9 326.8





#70
Styrene
Concen: 57.805 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:104 Resp: 16061
Ion Ratio Lower Upper
104 100
78 53.4 43.6 65.4
103 54.8 43.6 65.4


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#71
Bromoform
Concen: 51.837 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion:173 Resp: 43163
Ion Ratio Lower Upper
173 100
175 48.3 24.1 72.4
254 0.0 0.0 0.0





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:152 Resp: 6650
Ion Ratio Lower Upper
152 100
115 88.5 43.3 129.9
150 176.5 0.0 345.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#73
Isopropylbenzene
Concen: 54.069 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion:105 Resp: 253251
Ion Ratio Lower Upper
105 100
120 26.2 12.9 38.6





#74
N-amy1 acetate
Concen: 55.118 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD





Tgt Ion: 43 Resp: 11066

Ion	Ratio	Lower	Upper
43	100		
70	41.4	32.9	49.3
55	23.9	19.3	28.9
61	23.3	18.8	28.2


Manual Integrations
APPROVED

Reviewed By :John
Carlone



02/17/2024
Supervised By :Semsettin
Yesilyurt




02/20/2024




#75
1,1,2,2-Tetrachloroethane
Concen: 56.408 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57

Tgt Ion: 83 Resp: 94922


Ion	Ratio	Lower	Upper
83	100		
131	9.4	5.3	15.8
85	65.1	32.0	96.2





#76
1,2,3-Trichloropropane
Concen: 57.495 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 75 Resp: 7742
Ion Ratio Lower Upper
75 100
77 42.2 19.1 57.5


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#77
Bromobenzene
Concen: 57.013 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion:156 Resp: 63072
Ion Ratio Lower Upper
156 100
77 164.5 88.4 265.3
158 97.4 48.0 144.0





#78
n-propylbenzene
Concen: 54.654 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 91 Resp: 309923
Ion Ratio Lower Upper
91 100
120 22.4 11.5 34.4


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#79
2-Chlorotoluene
Concen: 53.608 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion: 91 Resp: 183568
Ion Ratio Lower Upper
91 100
126 33.3 16.3 48.8





#80
1,3,5-Trimethylbenzene
Concen: 54.753 ug/l
RT: 11.451 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:105 Resp: 215900
Ion Ratio Lower Upper
105 100
120 46.8 24.1 72.3


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#81
trans-1,4-Dichloro-2-butene
Concen: 49.074 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 75 Resp: 23500
Ion Ratio Lower Upper
75 100
53 106.9 90.5 135.7
89 47.2 36.9 55.3





#82
4-Chlorotoluene
Concen: 53.722 ug/l
RT: 11.451 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 91 Resp: 217960
Ion Ratio Lower Upper
91 100
126 28.9 14.2 42.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#83
tert-Butylbenzene
Concen: 54.298 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion:119 Resp: 209914
Ion Ratio Lower Upper
119 100
91 62.8 31.2 93.6
134 23.0 11.7 35.0





#84
1,2,4-Trimethylbenzene
Concen: 55.145 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:105 Resp: 219990
Ion Ratio Lower Upper
105 100
120 42.6 21.2 63.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#85
sec-Butylbenzene
Concen: 54.596 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion:105 Resp: 271255
Ion Ratio Lower Upper
105 100
134 19.8 9.8 29.4





#86
p-Isopropyltoluene
Concen: 54.745 ug/l
RT: 12.006 min Scan# 119
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:119 Resp: 22295
Ion Ratio Lower Upper
119 100
134 26.5 12.8 38.3
91 26.5 13.1 39.3


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#87
1,3-Dichlorobenzene
Concen: 54.771 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion:146 Resp: 120055
Ion Ratio Lower Upper
146 100
111 42.3 22.1 66.1
148 64.8 31.8 95.3





#88
1,4-Dichlorobenzene
Concen: 52.938 ug/l
RT: 12.042 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:146 Resp: 122872
Ion Ratio Lower Upper
146 100
111 40.9 21.3 63.7
148 63.1 31.8 95.4


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#89
n-Butylbenzene
Concen: 53.547 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion: 91 Resp: 210512
Ion Ratio Lower Upper
91 100
92 51.7 27.4 82.0
134 24.7 12.6 37.8





#90
Hexachloroethane
Concen: 57.935 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:117 Resp: 37218
Ion Ratio Lower Upper
117 100
201 73.4 32.9 98.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#91
1,2-Dichlorobenzene
Concen: 55.210 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion:146 Resp: 117152
Ion Ratio Lower Upper
146 100
111 43.6 22.4 67.0
148 64.1 31.6 95.0





#92
1,2-Dibromo-3-Chloropropane
Concen: 60.056 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.006 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion: 75 Resp: 2332
Ion Ratio Lower Upper
75 100
155 76.5 36.1 108.5
157 99.3 44.9 134.7


Manual Integrations
APPROVED

Reviewed By :John
Carlone




02/17/2024
Supervised By :Semsettin
Yesilyurt



02/20/2024




#93
1,2,4-Trichlorobenzene
Concen: 53.169 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57




Tgt Ion:180 Resp: 77240
Ion Ratio Lower Upper
180 100
182 93.7 48.3 144.8
145 31.3 18.0 54.0





#94
Hexachlorobutadiene
Concen: 49.678 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:225 Resp: 2880
Ion Ratio Lower Upper
225 100
223 63.7 31.1 93.2
227 63.3 32.5 97.5


Manual Integrations
APPROVED

Reviewed By :John
Carlone






02/17/2024
Supervised By :Semsettin
Yesilyurt


02/20/2024



#95
Naphthalene
Concen: 56.324 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57


Tgt Ion:128 Resp: 258428
Ion Ratio Lower Upper
128 100
127 12.9 10.3 15.5
129 11.0 8.8 13.2





#96
1,2,3-Trichlorobenzene
Concen: 54.450 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX040288.D
Acq: 16 Feb 2024 18:57



Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212MSD



Tgt Ion:180 Resp: 75398
Ion Ratio Lower Upper
180 100
182 95.6 46.9 140.6
145 33.1 17.8 53.5

Manual Integrations
APPROVED

Reviewed By :John
Carlone



02/17/2024
Supervised By :Semsettin
Yesilyurt

02/20/2024