



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021624\
Data File : VX040261.D
Acq On : 16 Feb 2024 08:13
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Title : SW846 8260
Last Update : Mon Feb 12 00:21:14 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	4781	PASS
75	95	30	60	53.5	13403	PASS
95	95	100	100	100.0	25035	PASS
96	95	5	9	7.2	1796	PASS
173	174	0.00	2	0.9	170	PASS
174	95	50	100	78.4	19634	PASS
175	174	5	9	7.5	1469	PASS
176	174	95	101	98.5	19345	PASS
177	176	5	9	6.4	1240	PASS