

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027069.D
 Acq On : 17 Feb 2022 11:36
 Operator : JC/MD
 Sample : VX0217WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0217WBL01

Quant Time: Feb 18 03:46:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	110345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	183866	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	165482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71977	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68715	51.877	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.760%
35) Dibromofluoromethane	5.391	113	57324	52.330	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.660%
50) Toluene-d8	8.653	98	213157	53.010	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.020%
62) 4-Bromofluorobenzene	11.079	95	78076	55.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027069.D
Acq On : 17 Feb 2022 11:36
Operator : JC/MD
Sample : VX0217WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0217WBL01

Quant Time: Feb 18 03:46:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

