

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027072.D
 Acq On : 17 Feb 2022 13:03
 Operator : JC/MD
 Sample : N1575-08DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-9RDL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/18/2022
 Supervised By : Mahesh Dadoda 02/21/2022

Quant Time: Feb 18 03:48:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	109346	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	179016	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163046	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	66718	50.829	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.660%			
35) Dibromofluoromethane	5.391	113	55918	52.429	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.860%			
50) Toluene-d8	8.653	98	208655	53.297	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 106.600%			
62) 4-Bromofluorobenzene	11.079	95	78725	57.116	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 114.240%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	1024m	1.776	ug/l	
31) Cyclohexane	5.483	56	8242	3.952	ug/l #	91
39) Methylcyclohexane	7.385	83	4701	2.278	ug/l	97
40) Benzene	6.038	78	3858	0.787	ug/l	100
52) Toluene	8.720	92	6707	2.160	ug/l	99
67) Ethyl Benzene	10.195	91	122736	20.566	ug/l	100
68) m/p-Xylenes	10.305	106	103334	45.003	ug/l	97
69) o-Xylene	10.646	106	24067	10.697	ug/l	98
73) Isopropylbenzene	10.963	105	16465	2.922	ug/l	97
78) n-propylbenzene	11.305	91	51563	8.254	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	41190	8.882	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	185370	39.628	ug/l	96
86) p-Isopropyltoluene	12.012	119	1572	0.331	ug/l #	72
95) Naphthalene	13.780	128	25227	4.819	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027072.D
Acq On : 17 Feb 2022 13:03
Operator : JC/MD
Sample : N1575-08DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-9RDL

Quant Time: Feb 18 03:48:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 02/18/2022
Supervised By : Mahesh Dadoda 02/21/2022

