

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027074.D
 Acq On : 17 Feb 2022 13:49
 Operator : JC/MD
 Sample : N1583-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 220216008-06-TRIP-BLANK

Quant Time: Feb 18 03:49:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	112926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	186844	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69488	51.261	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.520%	
35) Dibromofluoromethane	5.391	113	58481	52.535	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.080%	
50) Toluene-d8	8.653	98	214368	52.462	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.920%	
62) 4-Bromofluorobenzene	11.085	95	76675	53.298	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.600%	
Target Compounds						Qvalue
68) m/p-Xylenes	10.305	106	675	0.288	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027074.D
Acq On : 17 Feb 2022 13:49
Operator : JC/MD
Sample : N1583-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
220216008-06-TRIP-BLANK

Quant Time: Feb 18 03:49:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

