

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027075.D
 Acq On : 17 Feb 2022 14:13
 Operator : JC/MD
 Sample : N1575-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-24

Quant Time: Feb 18 03:49:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	108337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	183167	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	160144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68753	52.868	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.740%	
35) Dibromofluoromethane	5.391	113	57603	52.785	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.580%	
50) Toluene-d8	8.653	98	207780	51.870	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.740%	
62) 4-Bromofluorobenzene	11.085	95	74694	52.964	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.920%	
Target Compounds						Qvalue
16) Acetone	2.380	43	6841	11.977	ug/l	97
37) Ethyl Acetate	4.721	43	15722	8.036	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027075.D
Acq On : 17 Feb 2022 14:13
Operator : JC/MD
Sample : N1575-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-24

Quant Time: Feb 18 03:49:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

