

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027078.D
 Acq On : 17 Feb 2022 15:23
 Operator : JC/MD
 Sample : N1593-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11R

Quant Time: Feb 18 03:50:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/18/2022
 Supervised By : Mahesh Dadoda 02/21/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	110477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	187138	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	70238	52.963	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.920%			
35) Dibromofluoromethane	5.391	113	58690	52.640	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.280%			
50) Toluene-d8	8.653	98	215189	52.580	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.160%			
62) 4-Bromofluorobenzene	11.085	95	77008	53.446	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.900%			
Target Compounds						
16) Acetone	2.385	43	1878m	3.224	ug/l	
40) Benzene	6.043	78	5375	1.049	ug/l	100
73) Isopropylbenzene	10.963	105	2561	0.453	ug/l	100
85) sec-Butylbenzene	11.890	105	5078m	0.904	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027078.D
Acq On : 17 Feb 2022 15:23
Operator : JC/MD
Sample : N1593-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11R

Quant Time: Feb 18 03:50:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 02/18/2022
Supervised By : Mahesh Dadoda 02/21/2022

