

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027081.D
 Acq On : 17 Feb 2022 16:32
 Operator : JC/MD
 Sample : PB142725TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142725TB

Quant Time: Feb 18 03:51:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	110617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	188748	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	165333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	70583	53.156	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.320%			
35) Dibromofluoromethane	5.397	113	58753	52.247	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.500%			
50) Toluene-d8	8.653	98	216871	52.539	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.080%			
62) 4-Bromofluorobenzene	11.085	95	77256	53.160	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.320%			
Target Compounds						Qvalue
16) Acetone	2.380	43	6753	11.579	ug/l	94
20) Methylene Chloride	2.794	84	2566	2.031	ug/l	95
43) Isopropyl Acetate	6.348	43	11134	3.589	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027081.D
Acq On : 17 Feb 2022 16:32
Operator : JC/MD
Sample : PB142725TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142725TB

Quant Time: Feb 18 03:51:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

