

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027082.D
 Acq On : 17 Feb 2022 16:56
 Operator : JC/MD
 Sample : PB142725ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142725ZHE#01

Quant Time: Feb 18 03:52:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	104840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	178354	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66552	52.882	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.760%			
35) Dibromofluoromethane	5.391	113	55951	52.655	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.320%			
50) Toluene-d8	8.653	98	202855	52.007	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.020%			
62) 4-Bromofluorobenzene	11.079	95	73521	53.539	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 107.080%			
Target Compounds						
16) Acetone	2.386	43	6688	12.100	ug/l	98
20) Methylene Chloride	2.794	84	2509	2.095	ug/l	92
43) Isopropyl Acetate	6.348	43	10478	3.575	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027082.D
Acq On : 17 Feb 2022 16:56
Operator : JC/MD
Sample : PB142725ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142725ZHE#01

Quant Time: Feb 18 03:52:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

