

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027083.D
 Acq On : 17 Feb 2022 17:19
 Operator : JC/MD
 Sample : PB142725ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142725ZHE#02

Quant Time: Feb 18 03:52:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	107684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	183549	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	160144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68033	52.631	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.260%	
35) Dibromofluoromethane	5.391	113	57408	52.497	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.000%	
50) Toluene-d8	8.653	98	210226	52.372	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.740%	
62) 4-Bromofluorobenzene	11.085	95	73131	51.747	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.500%	
Target Compounds						Qvalue
16) Acetone	2.380	43	6717	11.831	ug/l	98
20) Methylene Chloride	2.794	84	2620	2.130	ug/l	87
43) Isopropyl Acetate	6.342	43	10962	3.634	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027083.D
Acq On : 17 Feb 2022 17:19
Operator : JC/MD
Sample : PB142725ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142725ZHE#02

Quant Time: Feb 18 03:52:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

