

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
 Data File : VX027089.D
 Acq On : 17 Feb 2022 19:39
 Operator : JC/MD
 Sample : N1561-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2-COMP

Quant Time: Feb 18 03:55:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	107055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	179242	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	67918	52.851	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.700%
35) Dibromofluoromethane	5.391	113	56437	52.849	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.700%
50) Toluene-d8	8.653	98	203952	52.030	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.060%
62) 4-Bromofluorobenzene	11.079	95	71985	52.160	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.320%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4808	8.518	ug/l	93
18) Methyl Acetate	2.703	43	736	0.453	ug/l	98
20) Methylene Chloride	2.794	84	1051	0.859	ug/l	96
43) Isopropyl Acetate	6.349	43	9460	3.211	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021722\
Data File : VX027089.D
Acq On : 17 Feb 2022 19:39
Operator : JC/MD
Sample : N1561-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2-COMP

Quant Time: Feb 18 03:55:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

